

North Inlet-Winyah Bay (NIW) NERR Water Quality Metadata
January 2007 – December 2007
Latest Update: 12/05/2010

I. Data Set and Research Descriptors

1) Principal investigator(s) and contact persons –

Address: North Inlet-Winyah Bay NERR
PO Box 1630
Georgetown, SC 29440
phone: (843)-546-3623
fax: (843)-546-1632

Contact Persons:

Dr. Erik Smith, Research Coordinator
email: erik@belle.baruch.sc.edu; (843)-546-3623 ext. 235
Tracy Buck, Research Specialist
email: tracy@belle.baruch.sc.edu; (843)-546-3623 ext. 259

2) Entry verification –

Deployment data are uploaded from the YSI data logger to a Personal Computer (IBM compatible). Files are exported from EcoWatch in a comma-delimited format (.CDF) and uploaded to the CDMO where they undergo automated primary QAQC and become part of the CDMO's online provisional database. Excessive pre- and post-deployment data are removed from the file prior to upload with up to 2 hours of pre- and post-deployment data retained to assist in data management. During primary QAQC, data are flagged if they are missing or out of sensor range. The edited file is then returned to the Reserve where it is opened in Microsoft Excel and processed using the CDMO's NERRQAQC Excel macro. The macro inserts station codes, creates metadata worksheets for flagged data, and graphs the data for review. It allows the user to apply QAQC flags and codes to the data, remove remaining pre- and post-deployment data, append files, and export the resulting data file to the CDMO for tertiary QAQC and assimilation into the CDMO's authoritative online database. Where deployment overlap occurs between files, the data produced by the newly calibrated sonde is generally accepted as being the most accurate. For more information on QAQC flags and codes, see Sections 11 and 12. Research Specialist, Tracy Buck edited and archived the 2007 data for Thousand Acre, Debidue Creek, Oyster Landing, and Clambank Creek.

3) Research objectives –

The principal objective of this study is to record long-term water quality data for North Inlet/Winyah Bay in order to observe any physical changes or trends in water quality over time. Four sites were chosen; two to represent pristine sites and two to represent impacted sites. The Oyster Landing site is located near the center of the Reserve and is considered pristine. The Clambank site is located on a waterway that receives run-off from the undeveloped area of the reserve and is considered a pristine site. Debidue Creek and Thousand Acre are both located in waterways that receive run-off from heavily developed areas, and are therefore considered our impacted sites. Measurements were taken every 15 minutes over roughly two week collecting periods at the Oyster Landing, Debidue Creek, Clambank and Thousand Acre sites.

4) Research methods –

YSI 6600 series sondes are currently mounted on either stand alone pilings (Debidue Creek, Oyster Landing) or bridge/dock pilings (Thousand Acre/Clambank Creek). Each sonde is deployed in a 4 inch diameter PVC pipe which is strapped to the piling using stainless steel banding. The PVC pipe is offset

from the pilings by inserting a 2x6 inch piece of treated lumber between the pipe and piling. This allows better flow over the sensors at the bottom of the pipes. The PVC pipe has a stainless steel stop bolt through its center 30 cm from the bottom of the pipe, ensuring that the sonde is seated 30 cm from the bottom of the creek. Directly above the stop bolt, 1 inch holes are drilled 0.5 cm apart within a 1 foot section of the pipe. This allows water flow over the sonde sensors. Four 1 inch holes are drilled perpendicular to each other at 1 foot intervals along the remaining length of the PVC pipe to allow for flushing. The bottom 3 feet of the pipe is painted inside and out with antifouling paint to minimize biofouling around the sensors. All pipes are switched out at least once a year (OL, CB, TA), while DC is switched out twice a year due to excessive fouling at that site.

Every 15 minutes measurements of specific conductivity, salinity, percent saturation, dissolved oxygen, water temperature, pH, turbidity, and water level are recorded. The two-week sampling interval was selected due to biofouling of the individual probes and expected battery life. Prior to deployment, the sondes are calibrated according to the procedures in the YSI Service Manual and the Standard Operating Procedures V4.1 provided by the NERR CDMO. Calibrations conducted prior to deployment of the instruments include a pH calibration using a two-point method with 7 and 10 unit standards. The turbidity calibration also uses a two-point method using 0 and 123 NTU standards. Specific Conductivity was calibrated using a one-point method with a 10 mS/cm standard that was purchased through YSI, Inc. Depth and dissolved oxygen are calibrated according to the barometric pressure at the time of calibration. Dissolved oxygen is calibrated using the air-saturated water method. The turbidity wiper and DO membrane of rapid-pulse DO probes were changed before each deployment. Beginning in 2007, ROX optical dissolved oxygen probes were installed on instruments at Debidue Creek, Thousand Acre and Oyster Landing, and run in tandem with traditional rapid-pulse probes with the intention of switching all sondes over to ROX probes by early 2008. At the end of each sampling interval, the sondes were brought back to the laboratory to be downloaded and cleaned. The sondes were allowed to continue recording data for at least 1 hour under lab conditions in a water-saturated air environment. Before the instruments were cleaned, a post calibration reading was taken from each instrument in fresh standard to see if any of the instruments exhibited drift (see Section 13). Post calibration readings of depth and dissolved oxygen are taken from the final readings of the unattended in-lab data. Two sondes are assigned to each of the four monitoring stations to allow the sondes to be switched out the same day, and minimize data loss.

A Sutron Sat-Link2 transmitter was installed at the Oyster Landing station on 07/07/06 and transmits data to the NOAA GOES satellite, NESDIS ID #3B031302. The transmissions are scheduled hourly and contain four (4) data sets reflecting fifteen minute data sampling intervals. Upon receipt by the CDMO, the data undergoes the same automated primary QAQC process detailed in Section 2 above. The “real-time” telemetry data become part of the provisional dataset until undergoing secondary and tertiary QAQC and assimilation in the CDMO’s authoritative online database. Provisional and authoritative data are available at <http://cdmo.baruch.sc.edu>.

5) Site location and character –

The North Inlet-Winyah Bay National Estuarine Research Reserve is located on the Southeastern Atlantic coast of the United States in two tidal estuaries, North Inlet and Winyah Bay, near Georgetown, South Carolina. The North Inlet estuary, located approximately 10 km east of Georgetown, is a bar-built Class C type estuary (Pritchard, 1955). The North Inlet estuary is composed of numerous winding tidal creeks, and is considered a pristine tidal estuary due to minimal anthropogenic impacts. The watershed drains a 24.8 km² area of mostly pine forest and a moderately developed residential watershed to the north. The Winyah Bay estuary, classified as a Class B type estuary by Pritchard (1955), which originates in the Blue Ridge Mountains of North Carolina, is one of the largest river-estuary ecosystems on the Eastern Seaboard. It is located 14.4 km south of North Inlet. Winyah Bay drains the sub-basins of 6 major rivers, which are heavily impacted by agriculture, mining, and industry. The rivers drain approximately 46,736 km² of uplands and marshes. Descriptions of the four sampling stations are as follows:

A) **Clambank Creek (CB)** - (lat 33:20:02, long 79:11:35). The Clambank Creek monitoring site is located roughly in the center of the reserve property. This site is surrounded by a *Spartina* marsh and drains associated uplands. Salinity can range from near fresh water concentrations following large rainfall events,

to full strength seawater during most of the year. The creek width and depth at MHW at the sampling site are 40 m and 2.4 m respectively. The average depth of the creek at MHW is 2.0 m with an average tidal range of approximately 1.4 meters. The bottom is comprised mostly of oyster shell hash and some fine sediment. This site is considered a pristine site and is influenced by its close proximity to North Inlet.

B) **Debidue Creek (DC)** - (lat 33:21:37, long 79:10:03). The Debidue Creek monitoring site is considered an impacted site that is located approximately 1 km south of the Debidue Colony. The Colony is a large development built on man-made canals that directly drain into the northern portion of Debidue Creek. The DC site is also located in an ocean-dominated *Spartina* marsh that was formerly surrounded by pine-dominated uplands. Salinity can range from 0 to full strength seawater and an average tidal range of approximately 2 m. The approximate depth and width at MHW at the site is 2.2 and 70 m respectively. The bottom is comprised mostly of oyster shell hash with some fine sediment and detritus.

C) **Oyster Landing (OL)** - (lat. 33:20:58, long. 79:11:20). The Oyster Landing monitoring site is considered a fairly pristine and undisturbed area located at the end of the Oyster Landing pier, which is also where the NI-WB Bay NERR weather station site is located. The pier stretches into the upper reaches of Crabhaul Creek in the mid western portion of North Inlet. The sampling site is located approximately 2.8 km from the headwaters of Crabhaul Creek. The creek directly drains pine-forested uplands and wetlands. Salinity can range from 0 to full strength seawater depending on rainfall events and upland drainage. The average tidal range is approximately 1.44 m. The creek has an average depth of 2 m MHW and average width of 150 m MHW. The bottom is comprised mostly of oyster shell hash with some fine sediment and detritus.

D) **Thousand Acre (TA)** - (lat. 33:17:57, long. 79:15:22) The Thousand Acre monitoring site is located in Thousand Acre marsh tidal creek and is on the NW corner of the west bridge of Thousand Acre marsh (this site was relocated 07-19-99 to the current location due to heavy siltation and degrading data quality). The present site is about 15 m from the mouth of the creek. At the sampling site, creek depth is approximately 2 m MHW and creek width is approximately 10 m. The creek empties into the northeastern side of the mid portion of Winyah Bay and directly drains pine forested upland and wetlands. Salinity ranges from 0 to 30 ppt. and tidal range is approximately 1m. The bottom is composed mostly of fine sediments and detritus. Georgetown, 5 km upstream from the Thousand Acre site and on the southern side of Winyah Bay, is the homeport for a number of heavy industries including a steel plant, paper mill, chemical plant, and a coal fired power plant. A public sewage treatment plant, which discharges into the bay, is also located in Georgetown.

6) Data collection period –

Thousand Acre data collection began January 1, 1995. Debidue Creek sampling began March 5, 1998. Oyster Landing data collection began in 1995; however, it was not considered a SWMP site until 1996 when the collection site was switched from Caledonia to Oyster Landing. Clambank sampling began in February 1981 through June 1995; however, it was not considered a SWMP site until sampling resumed on August 17, 2001. All sampling is ongoing.

Deployment dates and times (in Eastern Standard Time) for 2007 follow:

Site: Clambank Creek

Deploy Date	Time	Retrieve Date	Time
12/18/2006	10:15	01/05/2007	10:45
01/05/2007	11:00	01/18/2007	11:30
01/18/2007	11:45	02/05/2007	10:30
02/05/2007	10:45	02/20/2007	10:15
02/20/2007	10:30	03/06/2007	11:15
03/06/2007	11:30	03/20/2007	12:30
03/20/2007	12:45	04/04/2007	10:00

04/04/2007	10:15	04/17/2007	09:45
04/17/2007	10:00	05/03/2007	09:30
05/03/2007	09:45	05/18/2007	13:30
05/18/2007	13:45	05/31/2007	09:15
05/31/2007	09:30	06/19/2007	10:30
06/19/2007	10:45	07/05/2007	11:30
07/05/2007	11:45	07/23/2007	13:45
07/23/2007	14:00	07/31/2007	08:45
07/31/2007	09:00	08/17/2007	11:30
08/17/2007	11:45	08/30/2007	09:45
08/30/2007	10:00	09/12/2007	09:00
09/12/2007	09:15	09/28/2007	11:00
09/28/2007	11:30	10/12/2007	10:00
10/12/2007	10:15	10/29/2007	12:30
10/29/2007	12:45	11/13/2007	10:30
11/13/2007	10:45	11/27/2007	10:45
11/27/2007	11:00	12/12/2007	10:45
12/12/2007	11:00	01/10/2008	10:15

Site: Debidue Creek

Deploy Date	Time	Retrieve Date	Time
12/20/2006	09:15	01/05/2007	11:00
01/05/2007	11:15	01/18/2007	09:15
01/18/2007	09:30	02/05/2007	10:00
02/05/2007	10:15	02/20/2007	10:45
02/20/2007	11:00	03/06/2007	10:45*
03/06/2007	11:00	03/22/2007	09:30*
03/22/2007	09:45	04/04/2007	08:15*
04/04/2007	08:30	04/17/2007	09:15*
04/17/2007	09:30	05/03/2007	08:15
05/03/2007	08:30	05/17/2007	08:00*
05/17/2007	08:15	05/31/2007	08:00*
05/31/2007	08:15	06/19/2007	10:15
06/19/2007	10:30	07/05/2007	11:00*
07/05/2007	11:30	07/20/2007	11:30*
07/20/2007	11:45	07/31/2007	08:15*
07/31/2007	08:30	08/17/2007	10:45*
08/17/2007	11:00	08/30/2007	09:15*
08/30/2007	09:30	09/12/2007	08:30*
09/12/2007	08:45	09/28/2007	10:45*
09/28/2007	11:00	10/12/2007	09:30*
10/12/2007	09:45	10/29/2007	10:30*
10/29/2007	11:00	11/13/2007	10:15*
11/13/2007	10:30	11/27/2007	10:15*
11/27/2007	10:30	12/12/2007	10:30*
12/12/2007	10:45	01/01/2008	21:45*

*ROX Optical DO probes were used for the marked deployments

Site: Oyster Landing

Deploy Date	Time	Retrieve Date	Time
12/18/2006	09:45	01/05/2007	11:15
01/05/2007	11:30	01/19/2007	09:00
01/19/2007	09:15	02/06/2007	13:45

02/06/2007	14:00	02/08/2007	12:30
02/08/2007	15:00	02/20/2007	11:15
02/20/2007	11:30	02/20/2007	14:15**
02/20/2007	14:30	03/07/2007	10:30
03/07/2007	10:45	03/22/2007	10:45
03/22/2007	11:00	04/04/2007	10:30
04/04/2007	10:45	04/17/2007	10:00
04/17/2007	10:30	05/03/2007	10:00
05/03/2007	10:15	05/17/2007	10:30
05/17/2007	11:00	05/31/2007	09:30
05/31/2007	09:45	06/19/2007	12:00
06/19/2007	12:30	07/05/2007	12:30
07/05/2007	12:45	07/23/2007	14:00
07/23/2007	14:15	07/31/2007	09:45
07/31/2007	10:00	08/17/2007	12:00
08/17/2007	12:15	08/30/2007	11:00
08/30/2007	11:15	09/12/2007	10:15
09/12/2007	10:30	10/12/2007	11:00
10/12/2007	11:15	10/29/2007	11:00
10/29/2007	11:15	11/13/2007	11:45
11/13/2007	12:00	11/27/2007	11:00
11/27/2007	11:15	12/12/2007	11:15*
12/12/2007	11:30	01/10/2008	11:15*

*ROX Optical DO probes were used for the marked deployments

** The sonde malfunctioned and no data were reported for this deployment.

Site: **Thousand Acre**

Deploy Date	Time	Retrieve Date	Time
12/18/2006	11:00	01/05/2007	15:00
01/05/2007	15:15	01/18/2007	11:00
01/18/2007	11:15	02/05/2007	11:00
02/05/2007	11:15	02/21/2007	11:45
02/21/2007	12:00	03/06/2007	13:30
03/06/2007	13:45	03/20/2007	12:00
03/20/2007	12:15	04/04/2007	09:00*
04/04/2007	09:15	04/17/2007	11:15
04/17/2007	11:30	05/03/2007	09:00*
05/03/2007	09:15	05/18/2007	14:00
05/18/2007	14:15	05/31/2007	08:30*
05/31/2007	08:45	06/19/2007	11:15
06/19/2007	11:30	07/05/2007	12:00
07/05/2007	12:15	07/20/2007	12:15
07/20/2007	12:30	07/31/2007	09:15*
07/31/2007	09:30	08/17/2007	13:30
08/17/2007	13:45	08/30/2007	10:15*
08/30/2007	10:30	09/12/2007	09:30
09/12/2007	09:45	09/28/2007	11:30*
09/28/2007	11:45	10/12/2007	10:30*
10/12/2007	10:45	10/29/2007	08:30*
10/29/2007	08:45	11/13/2007	11:00*
11/13/2007	11:15	11/30/2007	10:45*
11/30/2007	11:00	12/14/2007	14:45
12/14/2007	15:00	01/10/2008	10:45

*ROX Optical DO probes were used for the marked deployments

7) Distribution –

NOAA/ERD retains the right to analyze, synthesize and publish summaries of the NERRS System-wide Monitoring Program data. The PI retains the right to be fully credited for having collected and processed the data. Following academic courtesy standards, the PI and NERR site where the data were collected will be contacted and fully acknowledged in any subsequent publications in which any part of the data are used. Manuscripts resulting from this NOAA/OCRM supported research that are produced for publication in open literature, including refereed scientific journals, will acknowledge that the research was conducted under an award from the Estuarine Reserves Division, Office of Ocean and Coastal Resource Management, National Ocean Service, National Oceanic and Atmospheric Administration. The data set enclosed within this package/transmission is only as good as the quality assurance and quality control procedures outlined by the enclosed metadata reporting statement. The user bears all responsibility for its subsequent use/misuse in any further analyses or comparisons. The Federal government does not assume liability to the Recipient or third persons, nor will the Federal government reimburse or indemnify the Recipient for its liability due to any losses resulting in any way from the use of this data.

NERR water quality data and metadata can be obtained from the Research Coordinator at the individual NERR site (please see Principal Investigators and Contact Persons), from the Data Manager at the Centralized Data Management Office (please see personnel directory under the general information link on the CDMO home page) and online at the CDMO home page <http://cdmo.baruch.sc.edu/>. Data are available in text tab-delimited format.

8) Associated researchers and projects

A water chemistry program is associated with the NIW core-monitoring program. Variables sampled include: chlorophyll, dissolved organic carbon, nitrate-nitrite, orthophosphate, and ammonia. (See documentation on the NIW WWW home page <http://www.northinlet.sc.edu/> for further details). Thousand Acre is our permanent monitoring station for the NERR monitoring program, but Oyster Landing and Debidue Creek were also sampled beginning in 1998. Clambank Landing was added as a sampling station in August of 2001. These stations are also included in our NIW core-monitoring program. Ben Lakish (Research Specialist) is responsible for the collection and management of this data.

The NERR weather station is also located at the Oyster Landing site. Air temperature and humidity, barometric pressure, solar radiation (total and PAR), wind speed and direction, and precipitation are measured. Amy Willman (Research Specialist) is responsible for the collection and management of this data.

In conjunction with NERR water quality and water chemistry sampling, Oyster Landing is the site of an ongoing study of microbial metabolism (production, respiration and bacterial growth efficiency) in salt marsh tidal creek waters. Sampling is done on both ebb and flood tides coincident with NIW NERR diel (ISCO) nutrient sampling. Study objectives include quantifying the factors that control the magnitude and variability in bacterioplankton metabolic rates and growth efficiencies, and the role of salt marsh exports (on tidal, seasonal, and interannual time scales) in fueling tidal creek heterotrophy. Dr. Erik Smith (NIW NERR Research Coordinator) is the principle investigator on this project.

A study of the optical characteristics of dissolved organic matter (DOM) is currently being conducted at Oyster Landing in association with the NIW NERR water quality and water chemistry sampling. Absorbance from 250 - 450 nm are quantified on filtered sub-samples of water collected during (ISCO) diel nutrient sampling and used to determine carbon-normalized DOM absorbance and spectral slope. Study objectives include using DOM optical characteristics as a tracer for the input of aromatic-rich DOM from higher plant and terrestrially-derived DOM sources to estuarine waters. Drs. Erik Smith (NIW NERR Research Coordinator) and Ralph Mead (Assistant Professor, UNCW) are the principle investigators on this project.

A long-term fish survey is also done at the Oyster Landing site bi-weekly. Species present, total weight, individual weights and standard lengths of fish that utilize the marsh at high tide are recorded. In the winter months (November through February), this sampling is done only once monthly. Dr. Dennis Allen (Director, Baruch Marine Lab) is the principal investigator on this project.

II. Physical Structure Descriptors

9) Sensor specifications –

NIWB NERR deployed both 6600 EDS and 6600 V2-4 sondes in 2007. Please note that v2-4 sondes do not have model 6562 Dissolved Oxygen probes (rapid pulse-clark cell), only model 6150 (ROX optical) probes. EDS sondes can run both 6562 and 6150 Dissolved Oxygen probes. All times below are in EST.

CB

Only EDS sondes with clark cell DO probes were deployed at CB during 2007.

DC

ROX probes were installed on EDS sondes and run in tandem with clark-cell probes at site DC from 02/05/2007 10:15 – 08/30/2007 9:30, and from 09/12/2007 8:45 – 10/29/2007 10:30. The data included in the DC data files for this period are from the rapid pulse probe for the 4/16 and 5/31 deployments, the ROX probe for the other deployments (this is also detailed in the Data Collection Period section). Clark cell probes were removed and only ROX probes were run from 10/29/2007 11:00 – 12/31/2007 23:30. A V2-4 sonde was deployed at DC from 08/30/2007 9:30 – 09/12/2007 08:30.

OL

EDS sondes with clark cell DO probes were deployed at OL from 01/01/2007 00:00 – 11/27/2007 11:00. V2-4 sondes were deployed at OL from 11/27/2007 11:15 – 12/31/2007 23:30.

TA

EDS sondes with clark cell DO probes were deployed at TA from 01/01/2007 00:00 – 03/20/2007 12:00, 04/04/2007 9:15 – 04/17/2007 11:15, 05/03/2007 9:15 – 05/18/2007 14:00, 05/31/2007 8:45 – 07/20/2007 12:15, 07/31/2007 9:30 – 08/17/2007 13:30, 08/30/2007 10:30 – 09/12/2007 09:30, and 11/30/2007 11:00 – 12/31/2007 23:30. All other times, a V2-4 sonde was deployed at TA.

YSI 6600 EDS-S Multi-parameter Water Quality Logger

Parameter:	Non-vented Level-Shallow (Depth)
Units:	feet or meters (ft or m)
Sensor Type:	Stainless steel strain gauge
Range:	0 to 30 ft (9.1 m)
Accuracy:	+/- 0.06 ft (0.018 m)
Resolution:	0.001 ft (0.001 m)

Parameter:	Temperature
Units:	Celsius (C)
Model #:	6560
Sensor Type:	Thermistor
Range:	-5 to 45 °C
Accuracy:	+/- 0.15 °C
Resolution:	0.01 °C

Parameter:	Dissolved Oxygen, Percent Saturation
Units:	percent air saturation (%)

Model #: 6562
 Sensor Type: Rapid Pulse - Clark type, polarographic
 Range: 0 to 500 % air saturation
 Accuracy: 0-200 % air saturation, +/- 2% of the reading or 2 % air saturation, whichever is greater, 200-500 % air saturation, +/- 6% of reading
 Resolution: 0.1 % air saturation

Parameter: Dissolved Oxygen, mg/L (calculated from % air saturation, temperature & salinity)
 Units: milligrams/Liter (mg/L)
 Model #: 6562
 Sensor Type: Rapid Pulse - Clark type, polarographic
 Range: 0 to 50 mg/L
 Accuracy: 0 to 20 mg/L, +/- 2% of the reading or 0.2 mg/L, whichever is greater, 20 to 50 mg/L, +/- 6% of the reading
 Resolution: 0.01 mg/L

Parameter: Dissolved Oxygen
 Model #: 6150
 Sensor Type: optical, self-cleaning
 Units: percent air saturation (%) and milligrams/Liter (mg/L)
 Range: 0 to 500 % air-saturation; 0 to 50 mg/L
 Accuracy: 0 to 200 % air-saturation, +/- 1 % of the reading or +/- 1 % air-saturation, whichever is greater; 200 to 500 % air-saturation, +/- 15 % of the reading 0 to 20 mg/L, +/- 1 % of the reading or +/- 0.1 mg/L, whichever is greater; 20 to 50 mg/L, +/- 15 % of the reading
 Resolution: 0.1 % air-saturation; 0.01 mg/L

Parameter: Conductivity
 Units: milli-Siemens per cm (mS/cm)
 Model #: 6560
 Sensor Type: 4 electrode cell with autoranging
 Range: 0 to 100 mS/cm
 Accuracy: +/- 0.5% of reading + 0.001 mS/cm
 Resolution: 0.001 mS/cm to 0.1 mS/cm (range dependent)

Parameter: Salinity
 Units: parts per thousand (ppt)
 Model #: 6560
 Sensor Type: Calculated from conductivity and temperature
 Range: 0 to 70 ppt
 Accuracy: +/- 1.0% of reading or 0.1 ppt, whichever is greater
 Resolution: 0.01 ppt

Parameter: pH
 Units: pH units
 Model #: 6561 or 6561FG (flat glass)
 Sensor Type: Glass combination electrode
 Range: 0 to 14 units
 Accuracy: +/- 0.2 units
 Resolution: 0.01 units

Parameter: Turbidity
 Units: nephelometric turbidity units (NTU)
 Model #: 6136
 Sensor Type: Optical, 90o scatter, with mechanical cleaning
 Range: 0 to 1000 NTU
 Accuracy: +/- 5% reading or 2 NTU (whichever is greater), relative to

Resolution: calibration standards
0.1 NTU

Dissolved Oxygen Qualifier:

The reliability of the dissolved oxygen (DO) data after 96 hours post-deployment for non-EDS (Extended Deployment System) data sondes may be problematic due to fouling which forms on the DO probe membrane during some deployments (Wenner et al. 2001). Many reserves have upgraded to the YSI 6600 EDS data sondes, which increases DO accuracy and longevity by reducing the environmental effects of fouling. The user is therefore advised to consult the metadata and to exercise caution when utilizing the DO data beyond the initial 96-hour time period. However, this potential drift is not always problematic for some uses of the data, i.e. periodicity analysis. It should also be noted that the amount of fouling is very site specific and that not all data are affected. The Research Coordinator at the specific NERR site should be contacted concerning the reliability of the DO data because of the site and seasonal variation in the fouling of the DO sensor.

Depth Qualifier:

The NERR System-Wide Monitoring Program utilizes YSI data sondes that can be equipped with either depth or water level sensors. Both sensors measure water depth, but by convention, level sensors refer to atmospherically vented measurements and depth refers to non-vented measurements. Readings for both vented and non-vented sensors are automatically compensated for water density change due to variations in temperature and salinity; but for all non-vented depth measurements, changes in atmospheric pressure between calibrations appear as changes in water depth. The error is equal to approximately 1.03 cm for every 1 millibar change in atmospheric pressure, and is eliminated for level sensors because they are vented to the atmosphere throughout the deployment time interval.

Beginning in 2006, NERR SWMP standard calibration protocol calls for all non-vented depth sensors to read 0 meters at a (local) barometric pressure of 1013.25 mb (760 mm/hg). To achieve this, each site calibrates their depth sensor with a depth offset number, which is calculated using the actual atmospheric pressure at the time of calibration and the equation provided in the SWMP calibration sheet or Digital Calibration Log. This offset procedure standardizes each depth calibration for the entire NERR System. If accurate atmospheric pressure data are available, non-vented sensor depth measurements at any NERR site can be corrected. The Research Coordinator at the specific NERR site should be contacted in order to obtain information regarding atmospheric pressure data availability.

10) Coded variable definitions –

Sampling station:	Sampling site code:	Station code:
Clambank Creek	CB	niwcbwq
Debidue Creek	DC	niwdcwq
Oyster Landing	OL	niwolwq
Thousand Acre	TA	niwtawq

11) QAQC flag definitions –

QAQC flags provide documentation of the data and are applied to individual data points by insertion into the parameter's associated flag column (header preceded by an F_). During primary automated QAQC (performed by the CDMO), -5, -4, and -2 flags are applied automatically to indicate data that is missing and above or below sensor range. All remaining data are then flagged 0, passing initial QAQC checks. During secondary and tertiary QAQC 1, -3, and 5 flags may be used to note data as suspect, rejected due to QAQC, or corrected.

- 5 Outside High Sensor Range
- 4 Outside Low Sensor Range
- 3 Data Rejected due to QAQC
- 2 Missing Data
- 1 Optional SWMP Supported Parameter
- 0 Data Passed Initial QAQC Checks
- 1 Suspect Data
- 2 *Open - reserved for later flag*
- 3 *Open - reserved for later flag*
- 4 Historical Data: Pre-Auto QAQC
- 5 Corrected Data

(NOTE: THIS IS A NEW SECTION)

12) QAQC code definitions –

QAQC codes are used in conjunction with QAQC flags to provide further documentation of the data and are also applied by insertion into the associated flag column. There are three (3) different code categories, general, sensor, and comment. General errors document general problems with the deployment or YSI datasonde, sensor errors are sensor specific, and comment codes are used to further document conditions or a problem with the data. Only one general or sensor error and one comment code can be applied to a particular data point.

General Errors

GIC	No Instrument Deployed Due to Ice
GIM	Instrument Malfunction
GIT	Instrument Recording Error; Recovered Telemetry Data
GMC	No Instrument Deployed Due to Maintenance/Calibration
GNF	Deployment Tube Clogged / No Flow
GOW	Out of Water Event
GPF	Power Failure / Low Battery
GQR	Data Rejected Due to QA/QC Checks
GSM	See Metadata

Sensor Errors

SBO	Blocked Optic
SCF	Conductivity Sensor Failure
SDF	Depth Port Frozen
SDO	DO Suspect
SDP	DO Membrane Puncture
SIC	Incorrect Calibration / Contaminated Standard
SNV	Negative Value
SOW	Sensor Out of Water
SPC	Post Calibration Out of Range
SSD	Sensor Drift
SSM	Sensor Malfunction
SSR	Sensor Removed / Not Deployed
STF	Catastrophic Temperature Sensor Failure
STS	Turbidity Spike
SWM	Wiper Malfunction / Loss

Comments

CAB	Algal Bloom
CAF	Acceptable Calibration/Accuracy Error of Sensor
CAP	Depth Sensor in Water, Affected by Atmospheric Pressure

CBF	Biofouling
CCU	Cause Unknown
CDA	DO Hypoxia (<3 mg/L)
CDB	Disturbed Bottom
CDF	Data Appear to Fit Conditions
CFK	Fish Kill
CIP	Surface Ice Present at Sample Station
CLT	Low Tide
CMC	In Field Maintenance/Cleaning
CMD	Mud in Probe Guard
CND	New Deployment Begins
CRE	Significant Rain Event
CSM	See Metadata
CTS	Turbidity Spike
CVT	Possible Vandalism/Tampering
CWD	Data Collected at Wrong Depth

13) Post deployment information –

Clambank Creek

PostCal Date	Sp.Con. (0 mS/cm)	Salinity (0 psu)	DO %	pH (7)	Depth	Turbidity (0 NTU)	BP (mm Hg)
01/05/2007	0.023		100.0	6.90	0.010	1.1	761.2
01/18/2007	0.003		100.2	6.94	0.110	0.2	768.3
02/05/2007	0.133	0.06	99.8	6.74	0.105	5.1	768.6
02/20/2007	0.063	0.03	101.8	6.97	0.057	-0.2	763.9
03/06/2007	0.064	0.03	103.0	7.12	0.125	0.1	769.8
03/20/2007	0.046	0.02	104.8		0.177	-0.4	773.4
04/04/2007	0.030	0.01	96.9	7.26	-0.043	0.1	756.1
04/17/2007	0.061	0.03	101.0	7.03	-0.082	0.0	753.5
05/03/2007	0.038	0.02	99.0	6.92	0.054	0.4	764.3
05/18/2007	0.033	0.01	97.0	6.99	0.013	1.2	760.9
05/31/2007	0.037	0.02	102.6	7.06	0.097	0.2	765.8
06/19/2007	0.026	0.01	98.3	7.12	0.036	0.7	762.8
07/05/2007	0.015	0.01	93.7	7.09	0.162	0.0	761.5
07/23/2007	0.038	0.02	100.1	6.98	-0.010	1.6	763.1
07/31/2007	0.035	0.02	99.8	7.03	-0.010	-1.6	758.6
08/17/2007			99.3		-0.023		760.7
08/30/2007	0.045	0.02	100.0	6.95	0.016	0.1	760.5
09/12/2007	0.032	0.01	98.4	6.98	0.007	0.0	761.4
09/28/2007			99.5		0.016		760.3
10/12/2007	0.075	0.03	105.1	6.78	0.041	0.6	759.1
10/29/2007	0.103	0.05	100.4	7.16	0.176	0.2	771.8
11/13/2007	0.087	0.04	99.6	6.67	0.002	0.2	762.5
11/27/2007	0.062	0.03	100.7	6.97	0.059	0.8	764.7
12/12/2007			94.3		0.043		765.3
01/10/2008	0.026	0.01	99.3	6.96	-0.002	0.5	759.1

Debidue Creek

PostCal Date	Sp. Cond. (0 mS/cm)	Salinity (0 psu)	DO % (% air sat.)	pH (7)	Depth	Turbidity (0 NTU)	BP (mm Hg)	ODO% (% air sat.)
01/05/2007	0.030		89.4	6.86	0.014	0.1	761.2	
01/18/2007	0.012		101.8	7.03	0.106	-0.1	767.8	
02/05/2007	0.028	0.01	100.9	7.03	0.110	0.3	768.7	

02/20/2007	0.116	0.05	99.9	6.90	0.052	0.4	763.9	
03/06/2007	0.047	0.02	102.7	6.88	0.125	0.3	769.8	104.8
03/22/2007	0.096	0.04	101.9	6.95	0.135	0.1	769.8	
04/04/2007	0.040	0.02	92.9	7.04	-0.056	0.0	755.7	
04/17/2007	0.059	0.03	100.6	7.11	-0.077	0.3	753.6	
05/03/2007	0.051	0.02	99.4	6.94	0.055	0.0	764.3	
05/17/2007	0.032	0.01	98.5	7.04	0.014	0.1	760.9	85.5
05/31/2007	0.046	0.02		7.12*	0.080	0.4	765.8	102.7
06/19/2007	0.015	0.01	104.7	6.95	0.033	0.1	762.8	
07/05/2007	0.175	0.08	101.9	6.91	0.057	-0.4	763.1	101.6
07/20/2007	0.028	0.01	98.5	7.02	-0.040	0.9	756.7	97.9
07/31/2007	0.021	0.01	98.5	6.95	-0.013	-0.7	758.6	
08/17/2007			103.1		0.008		760.6	99.8
08/30/2007	0.027	0.01	98.8	7.03	-0.014	0.0	758.5	97.9
09/12/2007	0.008	0.00		**	0.024	-0.3	761.4	102.5
09/28/2007			102.0		0.011		760.3	100.0
10/12/2007	0.009	0.00	100.2	7.02	-0.015	-0.3	759.1	101.2
10/29/2007	0.178	0.08	104.3	7.05	0.159	0.4	771.9	
11/13/2007	0.070	0.03		6.96	0.041	0.5	762.5	100.2
11/27/2007	0.064	0.03		6.97	0.059	-0.1	764.7	99.6
12/12/2007					0.081		765.3	99.8
01/10/2008	0.090	0.04		7.02	-0.018	0.0	759.6	***

* bubble in pH probe upon retrieval

** pH bulb broke upon retrieval

*** ROX probe failed during deployment

Oyster Landing

PostCal Date	Sp. Cond. (0 mS/cm)	Salinity (0 psu)	DO % (% air sat.)	pH (7)	Depth	Turbidity (0 NTU)	BP (mm Hg)	ODO% (% air sat.)
01/05/2007	0.007		101.2	6.77	0.016	-0.3	761.2	
01/19/2007	0.024		103.4	6.95	0.043	0.1		
02/06/2007	0.071	0.03	100.4	6.88	0.133	1.6	769.6	
02/20/2007	0.064	0.03	102.4	7.07	0.054	0.2	763.9	
03/07/2007	0.132	0.06	102.3	7.05	0.079	0.1	765.2	
03/22/2007	0.013	0.00	107.3	6.77	0.144	0.5	770.4	
04/04/2007	0.049	0.02	98.6	6.92	-0.060	0.7	755.7	
04/17/2007	0.035	0.02	98.0	7.08	-0.081	0.6	753.6	
05/03/2007	0.032	0.01	100.7	6.90	0.059	0.1	764.3	
05/17/2007	0.024	0.00	97.5	7.04	0.013	0.9	760.9	
05/31/2007	0.031	0.01	104.2	6.74	0.083	0.3	765.8	
06/19/2007	0.016	0.01	95.3	7.17	0.038	0.7	762.8	
07/05/2007	0.055	0.02	96.8	7.34	0.023	0.0	761.4	
07/23/2007	0.004	0.00	88.3*	6.97	0.049	0.1	763.1	
07/31/2007	0.035	0.01	99.0		0.001	0.9	758.6	
08/17/2007			111.6*		0.014		760.7	
08/30/2007	0.057	0.03	102.0	6.37	0.015	0.5	760.5	
09/12/2007	0.016	0.01	105.0	7.02	0.023	0.9	761.4	
10/12/2007	0.036	0.02	95.2	6.46	-0.008	0.9	759.1	
10/29/2007	0.049	0.02	105.4	6.92	0.165	0.9	771.9	
11/13/2007	0.065	0.03	99.3	7.00	0.030	0.4	762.7	
11/27/2007	0.059	0.03	98.9	6.73	0.062	0.6	764.7	
12/12/2007			100.5		0.072		765.3	
01/10/2008	0.042	0.02		7.05	-0.013	0.5	759.1	99.2

*DO membrane punctured

Thousand Acre

PostCal Date	Sp. Cond. (0 mS/cm)	Salinity (0 psu)	DO % (% air sat.)	pH (7)	Depth	Turbidity (0 NTU)	BP (mm Hg)	ODO% (% air sat.)
01/05/2007	0.001		100.5	6.84	0.009	-0.3	761.2	
01/18/2007	0.003*		103.2	6.94	0.111	0.0	768.6	
02/05/2007	0.018	0.01	86.3*	6.29	0.108	2.3	768.6	
02/21/2007	0.030	0.01	109.9	7.08	0.004	-0.2	759.9	
03/06/2007	0.028	0.01	110.5	7.00	0.127	-0.5	769.6	
03/20/2007	0.029	0.01	102.7	6.92	0.184	0.1	773.4	
04/04/2007	no post deployment data available							
04/17/2007	0.028	0.01	99.7	6.89	-0.085	0.1	753.6	
05/03/2007	0.026	0.01		6.98	0.057	-0.3	764.3	99.6
05/18/2007	0.038	0.02	99.6	6.89	0.004	0.4	760.9	
05/31/2007	0.000	0.00		7.05	0.081	0.5	765.8	94.1
06/19/2007	0.035	0.01	100.7	6.73	0.000	0.0	762.8	
07/05/2007	0.027	0.01	92.0	7.07	0.017	0.7	761.4	
07/20/2007	0.038	0.02	92.0	7.02	-0.036	0.2	756.7	
07/31/2007	0.055	0.02	97.4	7.08	-0.017	0.1	758.6	
08/17/2007			94.3**		0.012		760.7	
08/30/2007	0.068	0.03		7.12	0.011	0.9	760.5	98.2
09/12/2007	0.029	0.01	107.3	7.03	0.015	0.4	761.4	
09/28/2007					0.011		760.3	98.7
10/12/2007	0.036	0.02	104.9	7.13	-0.020	0.3	759.1	
10/29/2007	0.021	0.01		6.98	0.165	0.2	771.9	101.5
11/13/2007	0.073	0.03	99.2	7.11	0.033	0.2	762.7	
11/30/2007	0.025	0.01		7.11	0.104	-0.1	767.9	99.9
12/14/2007	no post deployment data available							
01/10/2008	0.011	0.00	99.6	6.87	0.027	0.1	762.0	

*temperature probe failure

**DO membrane punctured

14) Other remarks/notes –

- A) Data are missing due to equipment or associated specific probes not being deployed, equipment failure, time of maintenance or calibration of equipment, or repair/replacement of a sampling station platform. Any NaNs in the dataset stand for “not a number” and are the result of low power, disconnected wires, or out of range readings. If additional information on missing data is needed, contact the Research Coordinator at the reserve submitting the data.
- B) For the 8/17 12:15 – 8/30 11:00 deployment at Oyster Landing the sonde wasn’t pushed down all the way in the deployment tube during the deployment. Depth data are correct but flagged suspect because they do not reflect depth readings from the normal sonde location. All other parameters are coded CWD to indicate that they were collected at the wrong depth.
- C) For the 9/28 11:45 – 11/30 10:45 period at Thousand Acre, the sonde was not deployed at the corrected depth. Fouling within the tubing ripped the vexar mesh off the guards causing them to get jammed in the bottom of the tube. Technicians were unable to get the guards out so the sonde kept sitting up higher on top of them until they could switch out the pvc tube. Depth data are correct but flagged suspect because they do not reflect depth readings from the

normal sonde location. All other parameters are coded CWD to indicate that they were collected at the wrong depth.

- D) The following are daily rainfall totals > 2.54 mm for the year of 2007 recorded at the NIWB-NERR weather station at Oyster Landing. Note that significant rainfall amounts can affect all measured parameters, most noticeably salinity, turbidity and pH. Please note that Tropical events are associated with heavy winds that may also affect tidal depth.

Date	Total Daily Precip (mm)	
01/01/2007	02.8	
01/04/2007	05.6	
01/05/2007	08.1	
01/07/2007	04.1	
01/08/2007	03.3	
01/18/2007	13.5	
01/22/2007	27.7	
01/28/2007	14.5	
02/01/2007	28.2	
02/13/2007	12.7	
02/21/2007	12.4	
02/25/2007	05.6	
03/01/2007	05.3	
03/16/2007	04.1	
04/11/2007	14.7	
04/15/2007	06.1	
04/19/2007	06.1	
05/06/2007	03.6	
05/08/2007	08.1	
05/09/2007	26.2	Subtropical Storm Andrea
05/17/2007	19.6	
06/02/2007	46.5	
06/20/2007	14.2	
06/29/2007	12.7	
07/07/2007	11.2	
07/08/2007	05.1	
07/09/2007	03.6	
07/11/2007	06.1	
07/14/2007	16.0	
07/18/2007	05.8	
07/20/2007	11.4	
07/25/2007	42.7	
07/28/2007	19.8	
07/30/2007	09.4	
08/03/2007	20.1	
08/04/2007	12.2	
08/13/2007	07.4	
08/23/2007	04.6	
08/27/2007	08.9	
08/31/2007	18.3	

09/01/2007	03.6
09/12/2007	15.7
09/13/2007	01.3
09/14/2007	07.6
09/15/2007	10.4
09/20/2007	67.3

10/04/2007	21.1
10/18/2007	04.1
10/24/2007	06.6
10/25/2007	03.8
10/26/2007	14.0
10/27/2007	13.7

12/15/2007	06.4
12/16/2007	15.7
12/21/2007	28.2
12/25/2007	03.0
12/26/2007	07.9
12/29/2007	05.6
12/30/2007	03.0