## Old Woman Creek (OWC) NERR Site Water Quality Metadata

March through December, 2008 Latest Update: February 4, 2020

## 1. Principal Investigator & contact person:

Dr. David Klarer, Research Coordinator E-mail: david.klarer@oldwomancreek.org Old Woman Creek SNP and NERR 2514 Cleveland Road East Huron, Ohio 44839 Phone: (419) 433-4601

### 2. Entry verification:

Deployment data are uploaded from the YSI data logger to a Personal Computer (IBM compatible). Files are exported from EcoWatch in a comma-delimited format (.CDF) and uploaded to the CDMO where they undergo automated primary QAQC; automated depth/level corrections for changes in barometric pressure (cDepth or cLevel parameters); and become part of the CDMO's online provisional database. All pre- and post-deployment data are removed from the file prior to upload. During primary QAQC, data are flagged if they are missing or out of sensor range. The edited file is then returned to the Reserve for secondary QAQC where it is opened in Microsoft Excel and processed using the CDMO's NERRQAQC Excel macro. The macro inserts station codes, creates metadata worksheets for flagged data and summary statistics, and graphs the data for review. It allows the user to apply QAQC flags and codes to the data, remove any overlapping deployment data, append files, and export the resulting data file for upload to the CDMO. Upload after secondary QAQC results in ingestion into the database as provisional plus data, recalculation of cDepth or cLevel parameters, and finally tertiary QAQC by the CDMO and assimilation into the CDMO's authoritative online database. Where deployment overlap occurs between files, the data produced by the newly calibrated sonde is generally accepted as being the most accurate. For more information on QAQC flags and codes, see Sections 11 and 12.

The files are archived at OWC. Dr. David Klarer was responsible for both data logger deployment and data management at Old Woman Creek NERR during the 2007 deployment...

### 3. Research Objectives:

Measurements are taken every 15 minutes over two or three-week periods at four sites within the Old Woman Creek- three in the estuary proper- one in the upper reaches at State Route 2 (SU), one near the mouth, just south of State Route 6 (WM), and the third site upstream from the WM site (OL). The final site (BR) is just upstream of the first riffle zone above the estuary in Old Woman Creek proper. The purpose of this monitoring program is to document the role of this Great Lakes estuary in the Lake Erie ecosystem, particularly the estuary's role in mitigating storm flow that passes through it. In August Site SU had to be moved upstream approximately 0.3 kilometers to DR (Darrow Road), as the Ohio Department of Transportation was working on the bridge and the adjacent road at the SU site. The role of the OL site is to

document the degree of intrusion by lake water during northerly winds and subsequent seiche events.

#### 4. Research methods:

The YSI monitoring program began on 24 March, 2008 at all four sites. The sampling at all sites ended for the year on 3 December, 2008. Prior to deployment of the data loggers, a 4-inch diameter PVC pipe was bolted to an 8-foot long metal post that had been driven into the sediment. The logger trap at site DR was not bolted to an 8-foot metal post, but rather was suspended from the north side of the road bridge by metal chain. Each pipe had 4 rows of holes or 4 vertical slits 3/4" wide drilled into it spanning the area of the probe guard on the data logger to insure that the probes would have direct contact with the surrounding waters. Additional field readings for dissolved oxygen, pH, temperature, turbidity, and specific conductance are taken when the instrument is changed at each site (see the Other Remarks Section). The data loggers are replaced in the field after a two or three-week deployment, depending on temperature and degree of fouling of the data loggers. All data loggers were the extended deployment loggers. The data was retrieved from each data logger and each data logger was recalibrated (according to the directions in the YSI Operations Manual) before being returned to the field. Conductivity, turbidity (2 point calibration using distilled water for zero turbidity and a YSI standard for the other turbidity point), pH (2 point calibration), and turbidity (2 point calibration) are calibrated using commercial standards. These standards were prepared prior to each deployment. The data loggers at sites WM and SU have vented water level sensors while the loggers at sites BR and OL have non-vented depth sensors. At sites WM and intermittently at sites SU and DR, the ROX optical dissolved oxygen probe was used. The calibration logs provide sensor information. A Sutron Sat-Link2 transmitter was installed at Site OL during October 2006. This system transmits data to the NOAA Goes satellite, NESDIS ID# 3B02849A. The transmissions are scheduled hourly and contain four (4) data sets reflecting fifteen minute data sampling intervals. Upon receipt by the CDMO, the data undergoes the same automated primary QAQC process detailed in Section 2 above. The "real-time" telemetry data become part of the provisional dataset until undergoing secondary and tertiary QAOC and assimilation in the CDMO's authoritative online database. Provisional and authoritative data are available at http://cdmo.baruch.sc.edu.

### 5. Site Location and Character:

Old Woman Creek National Estuarine Research Reserve is located on the southern shore of Lake Erie, slightly east of the city of Huron, Ohio (Latitude 41° 23'N; Longitude 82° 33'W). Land use in the Old Woman Creek (OWC) watershed is primarily row crop agriculture. Other than the non-point source pollutants coming into the estuary from these agricultural practices and from the town of Berlin Heights, there are no other major pollution sources in the estuary. Salinity in Old Woman Creek is normally 1 ppt. or less, although it will rise, on occasion, to nearly 2 ppt. The tidal range in Lake Erie (and therefore in the estuary) is on the order of 4 cm or less. Water levels in the estuary and in the creek are extremely variable, with changes occurring daily, seasonally and annually due to changing lake levels, seiches on the lake, storm runoff, and the mouth closing and opening through the year.

The data logger at the State Route 6 (WM) site (Latitude 41° 22' 57" N, Longitude 82° 30'54" W) is very close to the mouth of Old Woman Creek. In this portion of the Reserve, the creek is very shallow but extends over a large surface area. This site frequently experiences influx of Lake Erie waters. The bottom sediments at this site are silty clay. At the beginning of

the deployment for 2007, there was no rooted aquatic vegetation directly adjacent to the site, although there was both emergent and submerged vegetation within 3 meters of the site. By mid September, *Phragmites australis* had grown up around the data logger trap. The data logger is about .18 meters above the bottom sediments.

The data logger at site OL (Latitude 41° 22' 55" N, Longitude 82° 30'51" W) is in the lower reaches of the estuary. This site is not in direct sight of the mouth, so northerly winds and resulting seiche activities should be less noticeable at this site. The bottom sediments are silty clay. This site is located about 5 meters north of a *Nelumbo lutea* bed, but, there were no plants immediately adjacent to the data logger. The base of data logger is about 31 cm above the sediment. This is the site that is telemetered to the GOES satellite.

The data logger at site BR (Latitude 41° 20'54" N, Longitude 82° 30'30"W) is located in the lower portion of the creek proper. Just upstream from the data logger, Berlin Road crosses Old Woman Creek. The site is just upstream of the first riffle area above the estuary. Unlike the other three sites, Lake Erie water levels have no impact on this site. The bottom of the creek at this site is a combination of rocks interspersed with some clay-silt that has been washed in from upstream. There are no aquatic macrophytes at or near this site. The trap is 18 cm above the bottom at this site. Short guards were used on all loggers at this site through the year. Wire mesh fencing around trap has been installed to diminish debris build-up around the logger. The trap was raised 10 cm on 5/26/2008 at 08:30. It was subsequently lowered 15 cm on 8/18/2008 at 08:15.

The data logger at site DR (Latitude 41<sup>o</sup> 21'54"N, Longitude 82<sup>o</sup> 30' 17"W) is at the southern boundary of the reserve. The logger trap is suspended from western most of the two center guard rail supports on the north side of the Darrow Road bridge near the deepest part of the creek channel. At this site the creek is relatively narrow. Although water direction and flow is influenced at this site by changes in Lake Erie water levels, this site doesn't have direct contact with Lake Erie waters. The bottom sediments at his site are silty clay. There is no rooted aquatic vegetation near or upstream from this site. The data logger is about .30 meters above the bottom at this site. The logger was move about 4 meters east and deployed in slightly deeper water and the trap was lowered 15 on 11/14/2008 at 09:30.

## 6. Data collection periods:

Sampling at WM began on March 24 2008 at 17:30. The logger was pulled for the year on December 3, 2008 at 12:15. Sampling at OL began on March 24, 2008 at 18:00, and ceased on December 3,2008 at 12:15. Sampling at BR began on March 25, 2008 at 07:30 and ceased on December 3, 2008 at 15:15. Sampling at DR began on March 24, 2008 at 08:15. The logger was pulled for the year on December 3, 2008 at 15:15. Specific deployment dates are listed below.

Site	Deployed	Pulled
WM	3/24/2008 (17:30)	4/14/2008 (07:15)
	4/14/2008 (07:30)	5/5/2008 (07:45)
	5/5/2008 (08:00)	5/26/2008 (08:45)
	5/26/2008 (08:00)	5/26/2008 (14:30)
	5/26/2008 (14:45)	6/16/2008 (07:30)
	6/16/2008 (07:45)	7/7/2008 (07:30)
	7/7/2008 (07:45)	7/29/2008 (08:00)
	7/29/2008 (08:15)	8/18/2008 (07:15)
	8/18/2008 (07:30)	9/8/2008 (07:15)
	9/8/2008 (07:30)	9/29/2008 (07:30)

	9/29/2008 (08:00) 10/11/2008 (09:15) 11/3/2008 (08:45) 11/24/2008 (08:30)	10/11/2008 (09:00) 11/3/2008 (08:30) 11/24/2008 (08:00) 12/3/2008 (12:15)
OL	3/24/2008 (18:00) 4/14/2008 (07:15) 5/5/2008 (08:15) 5/26/2008 (07:45) 6/16/2008 (07:30) 7/7/2008 (07:45) 7/29/2008 (08:15) 8/18/2008 (07:30) 9/8/2008 (07:15) 9/29/2008 (07:45) 10/11/2008 (09:00) 11/3/2008 (08:30) 11/26/2008 (09:15)	4/14/2008 (07:00) 5/5/2008 (08:00) 5/26/2008 (07:30) 6/16/2008 (07:15) 7/7/2008 (07:30) 7/29/2008 (08:00) 8/18/2008 (07:15) 9/8/2008 (07:00) 9/29/2008 (07:30) 10/11/2008 (08:45) 11/3/2008 (08:15) 11/26/2008 (09:00) 12/3/2008 (12:00)
BR	3/25/2008 (07:30) 4/14/2008 (08:15) 5/5/2008 (09:00) 5/26/2008 (08:45) 6/16/2008 (08:30) 7/7/2008 (08:30) 7/29/2008 (07:30) 8/18/2008 (08:30) 9/8/2008* (08:00) 9/29/2008 (08:30) 10/11/2008 (07:45) 11/3/2008 (09:30) 11/24/2008 (09:30)	4/14/2008 (08:00) 5/5/2008 (08:45) 5/26/2008 (08:30) 6/16/2008 (08:15) 7/7/2008 (08:15) 7/29/2008 (07:15) 8/18/2008 (07:15) 8/18/2008 (07:45) 9/29/2008* (08:15) 10/11/2008 (07:30) 11/3/2008 (09:15) 11/24/2008 (09:15) 12/3/2008 (15:15)
DR	3/24/2008 (08:15) 4/14/2008 (08:30) 5/5/2008 (09:15) 5/26/2008 (09:00) 6/16/2008 (08:45) 7/7/2008 (08:45) 7/29/2008 (07:45) 8/18/2008 (08:15) 9/8/2008 (08:15) 9/29/2008 (08:45) 10/11/2008 (08:15) 11/3/2008 (09:45) 11/24/2008 (09:45)	4/14/2008 (08:15) 5/5/2008 (09:00) 5/26/2008 (08:45) 6/16/2008 (08:15) 7/7/2008 (08:15) 7/29/2008 (07:30) 8/18/2008 (08:00) 9/8/2008 (08:00) 9/29/2008 (08:30) 10/11/2008 (08:00) 11/3/2008 (09:15) 11/24/2008 (09:30) 12/3/2008 (15:15)

\*There was a sonde in the water but No data was collected at BR during the 9/8 deployment due to a sonde malfunction.

#### 7. Distribution

NOAA/ERD retains the right to analyze, synthesize, and publish summaries of the NERRS System-wide Monitoring Program data. The OWC Research Coordinator (RC) retains the right to be fully credited for having collected and processed the data. Following academic courtesy standard, the RC and the NERR site where the data were collected will be contacted and fully acknowledged in any subsequent publications in which any part of the data are used. Manuscripts resulting from this NOAA/OCRM supported research that are produced for publication in open literature, including refereed scientific journals, will acknowledge that the research was conducted under an award from the Estuarine Reserves Division, Office of Ocean and Coastal Resource Management, National Ocean Service, National Oceanic and Atmospheric Administration.

The data set enclosed within this package/transmission is only as good as the quality assurance and quality control procedures outlined in the enclosed metadata reporting statement. The user bears all responsibility for its subsequent use/misuse in any further analyses or comparisons. The Federal government and the State of Ohio do not assume liability to the Recipient or third persons, nor will the Federal government or the State of Ohio reimburse or indemnify the Recipient for its liability due to any losses resulting in any way from the use of this data.

NERR water quality data and metadata can be obtained from the Research Coordinator at the individual NERR site (please see section 1, Principal Investigators and Contact Persons), from the Data Manager at the Centralized Data Management Office (please see personnel directory under general information link on CDMO homepage) and online at the CDMO homepage http://cdmo.baruch.sc.edu/. Data are available in text tab-delimited format.

### 8. Associated projects:

Replicate samples for chemical analysis of the water are collected at each site every time the data loggers are changed. Samples for phytoplankton determination are collected at the same time at sites near two of the data logger deployment sites (DR and WM). Additionally, a 26 hour sampling regime (samples are collected at 2 hour intervals over the 26 hours) is conducted at the WM site once during each month.

### **II. Physical Structure and Descriptors:**

#### 9. Sensor specifications:

YSI 6600EDS datalogger

Parameter: Temperature

Sensor Type: Thermistor

Model #: 6560 Range: -5 to 45 °C

Units: Celsius (C)

Accuracy: +/-0.15 °C Resolution: 0.01 °C

Parameter: Conductivity

Units: milli-Siemens per cm (mS/cm)

Sensor Type: 4-electrode cell with autoranging

Model #: 6560

Range: 0 to 100 mS/cm

Accuracy: +/-0.5% of reading + 0.001 mS/cm

Resolution: 0.001 mS/cm to 0.1 mS/cm (range dependent)

Parameter: Salinity

Units: parts per thousand (ppt)

Sensor Type: Calculated from conductivity and temperature

Range: 0 to 70 ppt

Accuracy: +/- 1.0% of reading or 0.1 ppt, whichever is greater

Resolution: 0.01 ppt

Parameter: Dissolved Oxygen % saturation

Units: percent air saturation (%)

Sensor Type: Rapid Pulse – Clark type, polarographic

Model #: 6562

Range: 0 to 500 % air saturation

Accuracy: 0-200 % air saturation, +/- 2 % of the reading or 2 % air saturation, whichever is

greater; 200-500 % air saturation, +/- 6 % of the reading

Resolution: 0.1 % air saturation

Parameter: Dissolved Oxygen mg/L (Calculated from % air saturation, temperature and salinity)

Units: milligrams per Liter (mg/L)

Sensor Type: Rapid Pulse – Clark type, polarographic

Model #: 6562 Range: 0 to 50 mg/L

Accuracy: 0 to 20 mg/L, +/- 2 % of the reading or 0.2 mg/L, whichever is greater; 20 to 50 mg/L,

+/- 6 % of the reading Resolution: 0.01 mg/L

Parameter: Dissolved Oxygen % saturation

Units: percent air saturation (%)

Sensor Type: Optical probe with mechanical cleaning

Model #: 6150 ROX

Range: 0 to 500% air saturation

Accuracy: 0-200% air saturation- +/- 1% of the reading or 1% air saturation, whichever is greater

200-500% air saturation- +/- 15% of the reading

Resolution: 0.1% air saturation

Parameter: Dissolved Oxygen mg/L (Calculated from % air saturation, temperature and salinity)

Units: milligrams per Liter (mg/L)

Sensor Type: Optical probe with mechanical cleaning

Model #: 6150 ROX Range: 0-50 mg/L

Accuracy: 0-20 mg/L- +/- 2% of the reading or 0.2 mg/L, whichever is greater

20-50 mg/L-  $\pm$  4% of the reading

Resolution: 0.01 mg/L

Parameter: Non-Vented Level – Shallow (Depth)

Units: feet or meters (ft or m)

Sensor Type: Stainless steel strain gauge

Range: 0 to 30 ft (9.1 m)

Accuracy: +/- 0.06 ft (0.018 m) Resolution: 0.001 ft (0.001 m)

Parameter: Vented Level – Shallow (Depth)

Units: feet or meters (ft or m)

Sensor Type: Stainless steel strain gauge

Range: 0 to 30 ft (9.1 m)

Accuracy 0-10 ft: +/- 0.01 ft (0.003 m) Accuracy 10-30 ft: +/- 0.06 ft (0.018 m)

Resolution: 0.001 ft (0.001 m)

Parameter: pH (EDS probe)

Units: units

Sensor Type: Glass combination electrode

Model #: 6561 Range: 0 to 14 units Accuracy: +/- 0.2 units

Resolution: 0.01 units

Parameter: Turbidity

Units: nephelometric turbidity units (NTU)

Sensor Type: Optical, 90 ° scatter, with mechanical cleaning

Model #: 6136

Range: 0 to 1000 NTU

Accuracy: +/- 5 % reading or 2 NTU (whichever is greater)

Resolution: 0.1 NTU

## **Depth Qualifier:**

The NERR System-Wide Monitoring Program utilizes YSI data sondes that can be equipped with either vented or non-vented depth/level sensors. Readings for both vented and non-vented sensors are automatically compensated for water density change due to variations in temperature and salinity; but for all nonvented depth measurements, changes in atmospheric pressure between calibrations appear as changes in water depth. The error is equal to approximately 1.03 cm for every 1 millibar change in atmospheric pressure, and is eliminated for vented sensors because they are vented to the atmosphere throughout the deployment time interval.

Beginning in 2006, NERR SWMP standard calibration protocol calls for all non-vented depth sensors to read 0 meters at a (local) barometric pressure of 1013.25 mb (760 mm/hg). To achieve this, each site calibrates their depth sensor with a depth offset number, which is calculated using the actual atmospheric pressure at the time of calibration and the equation provided in the SWMP calibration sheet or digital calibration log. This offset procedure standardizes each depth calibration for the entire NERR System. If accurate atmospheric pressure data are available, non-vented sensor depth measurements at any NERR can be corrected.

In 2010, the CDMO began automatically correcting depth/level data for changes in barometric pressure as measured by the Reserve's associated meteorological station during data ingestion. These corrected depth/level data are reported as cDepth and cLevel, and are assigned QAQC flags and codes based on QAQC protocols. Please see sections 11 and 12 for QAQC flag and code definitions.

#### 10. Coded variable definitions:

Sampling Station	Sampling site code	Station code
State Route 6	WM	owcwmwq
Lower Estuary	OL	owcolwq
Berlin Road	BR	owebrwq
Darrow Road	DR	owedrwq

## 11. QAQC flag

QAQC flags provide documentation of the data and are applied to individual data points by insertion into the parameter's associated flag column (header preceded by an F\_). During primary automated QAQC (performed by the CDMO), -5, -4, and -2 flags are applied automatically to indicate data that is missing and above or below sensor range. All remaining data are then flagged 0, passing initial QAQC checks. During secondary and tertiary QAQC 1, -3, and 5 flags may be used to note data as suspect, rejected due to QAQC, or corrected.

- -5 Outside High Sensor Range
- -4 Outside Low Sensor Range
- -3 Data Rejected due to QAQC
- -2 Missing Data
- -1 Optional SWMP Supported Parameter
- 0 Data Passed Initial QAQC Checks
- 1 Suspect Data
- 2 *Open reserved for later flag*
- 3 Calculated data: non-vented depth/level sensor correction for changes in barometric pressure
- 4 Historical Data: Pre-Auto QAQC
- 5 Corrected Data

### 12. QAQC code definitions

QAQC codes are used in conjunction with QAQC flags to provide further documentation of the data and are also applied by insertion into the associated flag column. There are three (3) different code categories, general, sensor, and

comment. General errors document general problems with the deployment or YSI datasonde, sensor errors are sensor specific, and comment codes are used to further document conditions or a problem with the data. Only one general or sensor error and one comment code can be applied to a particular data point, but some comment codes (marked with an \* below) can be applied to the entire record in the F\_Record column.

### **General Errors**

$\alpha$ T $\alpha$	3 T	•	1 1	1 1	
GIC	No	instrument	denlos	red due	to 100
OIC	110	mon union	ucbio v	cu uuc	to icc

GIM Instrument malfunction

GIT Instrument recording error; recovered telemetry data
GMC No instrument deployed due to maintenance/calibration

GNF Deployment tube clogged / no flow

GOW Out of water event

GPF Power failure / low battery

GQR Data rejected due to QA/QC checks

GSM See metadata

## Corrected Depth/Level Data Codes

GCC Calculated with data that were corrected during QA/QC

GCM Calculated value could not be determined due to missing data

GCR Calculated value could not be determined due to rejected data

GCS Calculated value suspect due to questionable data

GCU Calculated value could not be determined due to unavailable data

### Sensor Errors

SBO Blocked optic

SCF Conductivity sensor failure

SDF Depth port frozen

SDG Suspect due to sensor diagnostics

SDO DO suspect

SDP DO membrane puncture

SIC Incorrect calibration / contaminated standard

SNV Negative value

SOW Sensor out of water

SPC Post calibration out of range

SQR Data rejected due to QAQC checks

SSD Sensor drift

SSM Sensor malfunction

SSR Sensor removed / not deployed

STF Catastrophic temperature sensor failure

STS Turbidity spike

SWM Wiper malfunction / loss

#### Comments

CAB\* Algal bloom

CAF Acceptable calibration/accuracy error of sensor

CAP Depth sensor in water, affected by atmospheric pressure

CBF Biofouling

CCU Cause unknown

CDA\* DO hypoxia (<3 mg/L) CDB\* Disturbed bottom CDF Data appear to fit conditions CFK\* Fish kill CIP\* Surface ice present at sample station CLT\* Low tide CMC\* In field maintenance/cleaning Mud in probe guard CMD\* New deployment begins CND CRE\* Significant rain event See metadata CSM\* CTS Turbidity spike CVT\* Possible vandalism/tampering CWD\* Data collected at wrong depth CWE\* Significant weather event

## 13. Post deployment information:

End of Deployment Post-calibration Readings in Standard Solutions: Dissolved oxygen standard is in parentheses following the DO reading. Depth is always 0.0 meters for the vented loggers. For the unvented loggers, the depth reading in parentheses after the first depth reading is the expected depth reading when correcting for changes in barometric pressure. The specific conductivity standard is 1.413 mS/cm. The pH standard is 7.00. The primary turbidity standard is zero, and the second standard is in parentheses after the second turbidity reading. An asterisk after the DO% reading signifies that the DO membrane was punctured at time of retrieval. Complete post deployment data are in the calibration sheets

Site	Date	Sp. Cond.	DO(%)	pН	Turb	Depth
WM	4/14/2008	1.429	95.3	7.05	1.6; 123.9 (123)	001
	5/5/2008	1.421	96.9	7.03	1; 123.9 (123)	
	5/26/2008	1.429	96.9	7.04	0.9; 121.5 (121)	.001
	5/26/2008	logger failed			, , ,	
	6/16/2008	1.337	96.5	7.11	0.1; 110.8 (119)	.002
	7/7/2008	1.389	98.6	7.06	0.6; 121.6 (120)	.002
	7/29/2008	1.228	97.3	7.06	3.6; 123 (120)	001
	8/18/2008	1.387	90.4	7.03	0.4; 65.2 (120)	.003
	9/8/2008	1.213	97.5	7.05	2.6; 120.9 (120)	.002
	9/29/2008	1.430	98.6	7.11	4.1; 125.7 (121)	003
	10/11/2008	1.423	98.2	7.02	0.8; 123.2 (120)	.000
	11/3/2008	1.434	98.2	7.01	0.3; 112.1 (120)	005
	11/24/2008	1.407	97.3	6.93	1.3; 124.6 (120)	006
	12/3/2008	1.428	98.7	7.08	0.6; 119.5 (120)	004
	<sup>1</sup> non- v	ented logger us	ed during dep	oloyment		
Site	Date	Sp. Cond.	DO(%)	рН	Turb	Depth
OL	4/14/2008	1.419	97.0	7.05	0.4; 122.2 (123)	131 (135)

	5/5/2008	1.488	103.0	7.06	-0.2; 113.7 (123)	215 (203)
	5/26/2008	1.438	*	7.04	0.9; 122.4 (121)	` /
	6/16/2008	1.394	*	7.04	0.7; 124.4 (119)	
	7/7/2008	1.337	*	7.01	3.0; 123.8 (120)	
	7/29/2008	1.229	*	7.08	1.5; 122.8 (120)	, ,
	8/18/2008	1.271	*	7.03	1.6; 109.7 (120)	
	9/8/2008	1.265	*	6.88	2.7; 120.7 (120)	
	9/29/2008	1.388	103.7	6.88	0; 122.4 (121)	195 (185)
	10/11/2008	1.414	104.3	7.06	0.3; 121.2 (120)	` /
	11/3/2008	1.438	107.8	7.01	0.8; 122 (120)	152 (117)
	11/24/2008	1.504	*	7.34	1.0; 130.3 (120)	, ,
	12/3/2008	1.417	105.8	7.01	0.8; 121.4 (120)	031 (132)
				, , , ,	, ()	( )
Site	Date	Sp. Cond.	DO(%)	рН	Turb	Depth
BR	4/14/2008	1.371	98.5	7.02	0.6; 123 (123)	087 (097)
	5/5/2008	1.439	92.2	7.11	0.1; 123.7 (123)	197 (197)
	5/26/2008	1.417	95.5	7.02	1.5; 119.6 (121)	\ /
	6/16/2008	1.441	95.0	7.10	0.5; 123.5 (119)	285 (268)
	7/7/2008	1.428	100.6	7.07	0.9; 110.0 (120)	179 (173)
	7/29/2008	1.422	92.8	7.10	0.4; 122.0 (120)	184 (205)
	8/18/2008	1.381	93.9	6.99	0.1; 72.8 (120)	213 (205)
	9/8/2008	1.368	98.6	7.15	1.1; 119.8 (120)	137 (132)
	9/29/2008	logger faile	d during th	is deployment		
	10/11/2008	1.439	98.2	6.91	0.6; 120.8 (120)	066 (069)
	11/3/2008	1.395	96.8	6.97	0.1; 122.6 (120)	106 (095)
	11/24/2008	1.400	98.2	7.03	0.2; 122.1 (120)	297 (291)
	12/3/2008	1.419	98.7	6.92	0.1; 119.1 (120)	, ,
Site	Date	Sp. Cond.	DO(%)	pН	Turb	Depth
DR	4/14/2008	1.397	104.2	7.04	1.8; 124.3 (123)	127 (132)
	5/5/2008	1.452	74.3	7.04	0.3; 125.7 (123)	208 (198)
	5/26/2008	1.396	107.0	7.10	0; 121.7 (121)	285 (277)
	6/16/2008	1.434	98.3	6.96	1.8; 125.7 (119)	280 (266)
	7/7/2008	1.392	100.1	7.06	1.3; 125.1 (120)	188 (175)
	7/29/2008	1.413	100.5	7.09	1.2; 119.5 (120)	200 (211)
	8/18/2008	1.440	101.6	7.02	0.7; 122.9 (120)	217 (212)
	9/8/2008	1.322	114.8	7.15	1.3; 120.1 (120)	
	9/29/2008	1.433	106.3	6.94	0; 120.9 (120)	179 (182)
	10/11/2008	1.434	101.8	7.06	0.2; 120.9 (120)	
	11/3/2008	1.411	115.2	6.97	0.4; 123.1 (120)	` /
	11/24/2008	1.471	112.7	7.03	0.5; 124.5 (120)	
	12/3/2008	1.421	115.8	6.88	0.3; 120.3 (120)	
					, ( -)	` /

# 14. Other Remarks:

pH at BR was impacted by low water levels seasonally in 2008. Data that was impacted

by this issue have been marked 1 CSM throughout the year.

Depth data for site WM from 3/24/2008 (17:30) through 4/14/2008 (07:15) is rejected as the logger used was vented, but the vented cable to connect the logger was broken.

Data at for the WM site from 5/6 8:15 to 10:00 is marked -3 GSM CMC. The sonde was out of the tube while the venter line was reinstalled from the trap house.

Turbidity data at the WM site from 10/8 23:00 to 10/11 9:00 is marked suspect. There was a note in the field log about light debris that may have impacted turbidity data during this time.

SpCond and salinity data for the DR site from 11/21 12:30 to 16:00 appear to be impacted by low water, but are not all of the way out of the water. These data are marked suspect.

The conductivity post at OL for the 9/8 deployment was 0.388 in a 1.42 standard. After the sonde was cleaned the probe read ok. The data matches up well and appears ok despite the post reading very low.

The DO posts at WM for the last three deployments of the year (11/3, 11/24 and 12/3) were slightly odd. They read very low prior to cal and then over 112% at the post. The data matches up well and does not appear to be impacted. But, these posts were odd enough to discuss. No data has been marked suspect. More than likely DO readings during these periods were high due to low temperatures.

The trap for the WM site during the 11/24 deployment was lowered 15 cm prior to deployment. This caused a disjunct in the depth data compared to other deployments.

Field data collected at time of data logger swap is reported below. Specific conductivity was taken in the laboratory immediately after returning from the field. Temperature is reported in Degrees C, specific conductivity in millimhos, and oxygen in milligrams/liter.

Site	Date	Temp	Sp. Cond.	DO(mg/l)	pН
WM	3/24/2008	4.5	.478	11.80	7.80
	4/14/2008	7.1	.381	10.14	7.76
	5/5/2008	15.7	.511	5.30	7.68
	5/26/2008	18.7	.530	7.69	7.33
	5/26/2008	22.4		10.83	8.22
	6/16/2008	23.2	.638	5.91	7.72
	7/7/2008	22.8	.633	7.58	7.88
	7/29/2008	27.5	.337	6.29	8.01
	8/18/2008	23.8	.417	5.20	7.44
	9/8/2008	21.2	.433	4.68	7.69
	9/29/2008	20.9	.495	5.37	7.26
	10/11/2008	17.4	.511	5.89	7.47
	11/3/2008	10.6	.509	9.89	7.81

Site   Date   Temp   Sp. Cond.   DO(mg/l)   pH		11/24/2008	1.9	.756	11.66	7.77
OL 3/24/2008 5.3 .478 11.85 7.92 4/14/2008 7.0 .483 10.58 7.89 5/5/2008 15.1 .477 8.45 7.75 5/26/2008 19.3 .630 6.65 7.74 6/16/2008 23.0 .701 6.11 7.65 7/7/2008 23.8 .568 7.14 7.74 7/29/2008 26.9 .356 5.76 7.88 8/18/2008 23.4 .419 4.83 7.52 9/8/2008 20.5 .432 4.88 7.59 9/29/2008 20.8 .488 4.88 7.25 10/11/2008 17.8 .508 5.94 7.48 11/3/2008 10.8 .509 9.75 7.69 11/26/2008 3.2 .804 11.85 7.84 12/3/2008 4.5 .598 11.70 7.68  Site Date Temp Sp. Cond. DO(mg/l) pH  BR 3/24/2008 2.0 .489 11.8 7.80 5/5/2008 13.9 .513 9.57 7.89 5/26/2008 13.9 .513 9.57 7.89 5/26/2008 13.9 .513 9.57 7.89 5/26/2008 13.9 .513 9.57 7.89 5/26/2008 22.1 .761 6.41 7.78 8/18/2008 22.1 .761 6.41 7.78 8/18/2008 20.4 .864 4.96 7.42 9/8/2008 18.2 .976 5.39 7.60 9/29/2008 18.9 .769 5.20 7.53 10/11/2008 18.9 .769 5.20 7.53 10/11/2008 18.9 .769 5.20 7.53 10/11/2008 18.9 .769 5.20 7.53 11/3/2008 18.9 .769 5.20 7.53 11/3/2008 18.9 .769 5.20 7.53 11/3/2008 2.4 .746 12.98 8.07 11/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 2.4 .746 12.98 8.07 11/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 2.4 .746 12.98 8.07 11/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 2.4 .746 12.98 8.07 11/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 3.5 .438 12.10 7.86 4/14/2008 7.3 .509 10.92 7.86 5/5/2008 15.2 .502 8.84 7.86 5/5/2008 15.2 .502 8.84 7.86 5/5/2008 15.2 .502 8.84 7.86 5/5/2008 15.2 .502 8.84 7.86 5/5/2008 23.9 .753 6.86 7.90 8/18/2008 23.9 .753 6.86 7.90 8/18/2008 23.9 .753 6.86 7.90 8/18/2008 23.9 .753 6.86 7.90 8/18/2008 22.10 .747 5.27 7.86						
OL 3/24/2008 5.3 .478 11.85 7.92 4/14/2008 7.0 .483 10.58 7.89 5/5/2008 15.1 .477 8.45 7.75 5/26/2008 19.3 .630 6.65 7.74 6/16/2008 23.0 .701 6.11 7.65 7/7/2008 23.8 .568 7.14 7.74 7/29/2008 26.9 .356 5.76 7.88 8/18/2008 23.4 .419 4.83 7.52 9/8/2008 20.5 .432 4.88 7.59 9/29/2008 20.8 .488 4.88 7.25 10/11/2008 17.8 .508 5.94 7.48 11/3/2008 10.8 .509 9.75 7.69 11/26/2008 3.2 .804 11.85 7.84 12/3/2008 4.5 .598 11.70 7.68  Site Date Temp Sp. Cond. DO(mg/l) pH  BR 3/24/2008 2.0 .489 11.8 7.80 5/5/2008 13.9 .513 9.57 7.89 5/26/2008 13.9 .513 9.57 7.89 5/26/2008 13.9 .513 9.57 7.89 5/26/2008 13.9 .513 9.57 7.89 5/26/2008 22.1 .761 6.41 7.78 8/18/2008 22.1 .761 6.41 7.78 8/18/2008 20.4 .864 4.96 7.42 9/8/2008 18.2 .976 5.39 7.60 9/29/2008 18.9 .769 5.20 7.53 10/11/2008 18.9 .769 5.20 7.53 10/11/2008 18.9 .769 5.20 7.53 10/11/2008 18.9 .769 5.20 7.53 11/3/2008 18.9 .769 5.20 7.53 11/3/2008 18.9 .769 5.20 7.53 11/3/2008 2.4 .746 12.98 8.07 11/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 2.4 .746 12.98 8.07 11/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 2.4 .746 12.98 8.07 11/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 2.4 .746 12.98 8.07 11/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 3.5 .438 12.10 7.86 4/14/2008 7.3 .509 10.92 7.86 5/5/2008 15.2 .502 8.84 7.86 5/5/2008 15.2 .502 8.84 7.86 5/5/2008 15.2 .502 8.84 7.86 5/5/2008 15.2 .502 8.84 7.86 5/5/2008 23.9 .753 6.86 7.90 8/18/2008 23.9 .753 6.86 7.90 8/18/2008 23.9 .753 6.86 7.90 8/18/2008 23.9 .753 6.86 7.90 8/18/2008 22.10 .747 5.27 7.86		_	_			
4/14/2008	Site	Date	Temp	Sp. Cond.	DO(mg/l)	pН
4/14/2008	OL	3/24/2008	5.3	.478	11.85	7.92
5/5/2008         15.1         .477         8.45         7.75           5/26/2008         19.3         .630         6.65         7.74           6/16/2008         23.0         .701         6.11         7.65           7/7/2008         23.8         .568         7.14         7.74           7/29/2008         26.9         .356         5.76         7.88           8/18/2008         23.4         .419         4.83         7.52           9/8/2008         20.5         .432         4.48         7.59           9/29/2008         20.8         .488         4.88         7.25           10/11/2008         17.8         .508         5.94         7.48           11/3/2008         10.8         .509         9.75         7.69           11/26/2008         3.2         .804         11.85         7.84           12/3/2008         4.5         .598         11.70         7.68           Site         Date         Temp         Sp. Cond.         DO(mg/l)         pH           BR         3/24/2008         2.0         .489         11.8         7.80           5/5/2008         13.9         .513         9.57         7.89	02					
5/26/2008         19.3         .630         6.65         7.74           6/16/2008         23.0         .701         6.11         7.65           7/7/2008         23.8         .568         7.14         7.74           7/29/2008         26.9         .356         5.76         7.88           8/18/2008         23.4         .419         4.83         7.52           9/8/2008         20.5         .432         4.48         7.59           9/29/2008         20.8         .488         4.88         7.25           10/11/2008         17.8         .508         5.94         7.48           11/3/2008         10.8         .509         9.75         7.69           11/26/2008         3.2         .804         11.85         7.84           12/3/2008         4.5         .598         11.70         7.68           Site         Date         Temp         Sp. Cond.         DO(mg/l)         pH           BR         3/24/2008         2.0         .489         11.8         7.80           4/14/2008         6.9         .505         11.34         7.94           5/5/2008         13.9         .513         9.57         7.89						
6/16/2008						
7/7/2008						
7/29/2008 26.9 .356 5.76 7.88 8/18/2008 23.4 .419 4.83 7.52 9/8/2008 20.5 .432 4.48 7.59 9/29/2008 20.8 .488 4.88 7.25 10/11/2008 17.8 .508 5.94 7.48 11/3/2008 10.8 .509 9.75 7.69 11/26/2008 3.2 .804 11.85 7.84 12/3/2008 4.5 .598 11.70 7.68  Site Date Temp Sp. Cond. DO(mg/l) pH  BR 3/24/2008 2.0 .489 11.8 7.80 4/14/2008 6.9 .505 11.34 7.94 5/5/2008 13.9 .513 9.57 7.89 5/26/2008 19.7 .673 8.20 7.90 6/16/2008 21.8 .752 6.95 7.82 7/7/2008 22.8 .633 7.58 7.88 7/29/2008 22.1 .761 6.41 7.78 8/18/2008 20.4 .864 4.96 7.42 9/8/2008 18.2 .976 5.39 7.60 9/8/2008 18.5 .947 6.75 7.03 11/3/2008 13.5 .947 6.75 7.03 11/3/2008 13.5 .947 6.75 7.03 11/3/2008 2.4 .769 5.20 7.53 10/11/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 2.4 .746 12.98 8.07 12/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 3.5 .438 12.10 7.86 4/14/2008 7.3 .509 10.92 7.86 5/5/2008 15.2 .502 8.84 7.86 5/26/2008 17.9 .677 7.17 7.79 6/16/2008 21.8 .766 6.93 7.75 7/7/2008 23.4 .606 7.44 7.88 7/29/2008 23.9 .753 6.86 7.90 8/18/2008 23.9 .753 6.86 7.90 8/18/2008 22.9 .737 6.12 7.73 9/8/2008 21.0 .747 5.27 7.86						
8/18/2008						
9/8/2008						
9/29/2008						
10/11/2008 17.8 .508 5.94 7.48 11/3/2008 10.8 .509 9.75 7.69 11/26/2008 3.2 .804 11.85 7.84 12/3/2008 4.5 .598 11.70 7.68  Site Date Temp Sp. Cond. DO(mg/l) pH  BR 3/24/2008 2.0 .489 11.8 7.80 4/14/2008 6.9 .505 11.34 7.94 5/5/2008 13.9 .513 9.57 7.89 5/26/2008 19.7 .673 8.20 7.90 6/16/2008 21.8 .752 6.95 7.82 7/7/2008 22.8 .633 7.58 7.88 7/29/2008 22.1 .761 6.41 7.78 8/18/2008 20.4 .864 4.96 7.42 9/8/2008 18.2 .976 5.39 7.60 9/29/2008 18.9 .769 5.20 7.53 10/11/2008 13.5 .947 6.75 7.03 11/3/2008 10.4 .855 8.06 7.68 11/24/2008 2.4 .746 12.98 8.07 12/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 3.5 .438 12.10 7.86 4/14/2008 7.3 .509 10.92 7.86 5/5/2008 15.2 .502 8.84 7.86 5/26/2008 17.9 .677 7.17 7.79 6/16/2008 21.8 .766 6.93 7.75 7/7/2008 23.4 .606 7.44 7.88 7/29/2008 23.9 .753 6.86 7.90 8/18/2008 22.9 .737 6.12 7.73 9/8/2008 22.9 .737 6.12 7.73		9/29/2008		.488		7.25
11/26/2008         3.2         .804         11.85         7.84           12/3/2008         4.5         .598         11.70         7.68           Site         Date         Temp         Sp. Cond.         DO(mg/l)         pH           BR         3/24/2008         2.0         .489         11.8         7.80           4/14/2008         6.9         .505         11.34         7.94           5/5/2008         13.9         .513         9.57         7.89           5/26/2008         19.7         .673         8.20         7.90           6/16/2008         21.8         .752         6.95         7.82           7/7/2008         22.8         .633         7.58         7.88           7/29/2008         22.1         .761         6.41         7.78           8/18/2008         20.4         .864         4.96         7.42           9/8/2008         18.2         .976         5.39         7.60           9/29/2008         18.9         .769         5.20         7.53           10/11/2008         13.5         .947         6.75         7.03           11/3/2008         10.4         .855         8.06         7.68			17.8		5.94	
11/26/2008         3.2         .804         11.85         7.84           12/3/2008         4.5         .598         11.70         7.68           Site         Date         Temp         Sp. Cond.         DO(mg/l)         pH           BR         3/24/2008         2.0         .489         11.8         7.80           4/14/2008         6.9         .505         11.34         7.94           5/5/2008         13.9         .513         9.57         7.89           5/26/2008         19.7         .673         8.20         7.90           6/16/2008         21.8         .752         6.95         7.82           7/7/2008         22.8         .633         7.58         7.88           7/29/2008         22.1         .761         6.41         7.78           8/18/2008         20.4         .864         4.96         7.42           9/8/2008         18.2         .976         5.39         7.60           9/29/2008         18.9         .769         5.20         7.53           10/11/2008         13.5         .947         6.75         7.03           11/3/2008         10.4         .855         8.06         7.68			10.8			
Site         Date         Temp         Sp. Cond.         DO(mg/l)         pH           BR         3/24/2008         2.0         .489         11.8         7.80           4/14/2008         6.9         .505         11.34         7.94           5/5/2008         13.9         .513         9.57         7.89           5/26/2008         19.7         .673         8.20         7.90           6/16/2008         21.8         .752         6.95         7.82           7/7/2008         22.8         .633         7.58         7.88           7/29/2008         22.1         .761         6.41         7.78           8/18/2008         20.4         .864         4.96         7.42           9/8/2008         18.2         .976         5.39         7.60           9/29/2008         18.9         .769         5.20         7.53           10/11/2008         13.5         .947         6.75         7.03           11/3/2008         10.4         .855         8.06         7.68           11/24/2008         2.4         .746         12.98         8.07           12/3/2008         2.8         .633         12.12         7.94		11/26/2008	3.2	.804	11.85	7.84
BR 3/24/2008 2.0 .489 11.8 7.80 4/14/2008 6.9 .505 11.34 7.94 5/5/2008 13.9 .513 9.57 7.89 5/26/2008 19.7 .673 8.20 7.90 6/16/2008 21.8 .752 6.95 7.82 7/7/2008 22.8 .633 7.58 7.88 7/29/2008 22.1 .761 6.41 7.78 8/18/2008 20.4 .864 4.96 7.42 9/8/2008 18.2 .976 5.39 7.60 9/29/2008 18.9 .769 5.20 7.53 10/11/2008 13.5 .947 6.75 7.03 11/3/2008 10.4 .855 8.06 7.68 11/24/2008 2.4 .746 12.98 8.07 12/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 3.5 .438 12.10 7.86 4/14/2008 7.3 .509 10.92 7.86 5/5/2008 15.2 .502 8.84 7.86 5/26/2008 17.9 .677 7.17 7.79 6/16/2008 23.4 .766 6.93 7.75 7/7/2008 23.4 .606 7.44 7.88 7/29/2008 23.9 .753 6.86 7.90 8/18/2008 22.9 .737 6.12 7.73 9/8/2008 22.9 .737 6.12 7.73 9/8/2008 21.0 .747 5.27 7.86		12/3/2008	4.5	.598	11.70	7.68
BR 3/24/2008 2.0 .489 11.8 7.80 4/14/2008 6.9 .505 11.34 7.94 5/5/2008 13.9 .513 9.57 7.89 5/26/2008 19.7 .673 8.20 7.90 6/16/2008 21.8 .752 6.95 7.82 7/7/2008 22.8 .633 7.58 7.88 7/29/2008 22.1 .761 6.41 7.78 8/18/2008 20.4 .864 4.96 7.42 9/8/2008 18.2 .976 5.39 7.60 9/29/2008 18.9 .769 5.20 7.53 10/11/2008 13.5 .947 6.75 7.03 11/3/2008 10.4 .855 8.06 7.68 11/24/2008 2.4 .746 12.98 8.07 12/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 3.5 .438 12.10 7.86 4/14/2008 7.3 .509 10.92 7.86 5/5/2008 15.2 .502 8.84 7.86 5/26/2008 17.9 .677 7.17 7.79 6/16/2008 23.4 .766 6.93 7.75 7/7/2008 23.4 .606 7.44 7.88 7/29/2008 23.9 .753 6.86 7.90 8/18/2008 22.9 .737 6.12 7.73 9/8/2008 22.9 .737 6.12 7.73 9/8/2008 21.0 .747 5.27 7.86		_				
4/14/2008       6.9       .505       11.34       7.94         5/5/2008       13.9       .513       9.57       7.89         5/26/2008       19.7       .673       8.20       7.90         6/16/2008       21.8       .752       6.95       7.82         7/7/2008       22.8       .633       7.58       7.88         7/29/2008       22.1       .761       6.41       7.78         8/18/2008       20.4       .864       4.96       7.42         9/8/2008       18.2       .976       5.39       7.60         9/29/2008       18.9       .769       5.20       7.53         10/11/2008       13.5       .947       6.75       7.03         11/3/2008       10.4       .855       8.06       7.68         11/24/2008       2.4       .746       12.98       8.07         12/3/2008       2.8       .633       12.12       7.94         Site       Date       Temp       Sp. Cond.       DO(mg/l)       pH         DR       3/24/2008       3.5       .438       12.10       7.86         5/5/2008       15.2       .502       8.84       7.86	Site	Date	Temp	Sp. Cond.	DO(mg/l)	рН
4/14/2008       6.9       .505       11.34       7.94         5/5/2008       13.9       .513       9.57       7.89         5/26/2008       19.7       .673       8.20       7.90         6/16/2008       21.8       .752       6.95       7.82         7/7/2008       22.8       .633       7.58       7.88         7/29/2008       22.1       .761       6.41       7.78         8/18/2008       20.4       .864       4.96       7.42         9/8/2008       18.2       .976       5.39       7.60         9/29/2008       18.9       .769       5.20       7.53         10/11/2008       13.5       .947       6.75       7.03         11/3/2008       10.4       .855       8.06       7.68         11/24/2008       2.4       .746       12.98       8.07         12/3/2008       2.8       .633       12.12       7.94         Site       Date       Temp       Sp. Cond.       DO(mg/l)       pH         DR       3/24/2008       3.5       .438       12.10       7.86         5/5/2008       15.2       .502       8.84       7.86	BR	3/24/2008	2.0	.489	11.8	7.80
5/5/2008         13.9         .513         9.57         7.89           5/26/2008         19.7         .673         8.20         7.90           6/16/2008         21.8         .752         6.95         7.82           7/7/2008         22.8         .633         7.58         7.88           7/29/2008         22.1         .761         6.41         7.78           8/18/2008         20.4         .864         4.96         7.42           9/8/2008         18.2         .976         5.39         7.60           9/29/2008         18.9         .769         5.20         7.53           10/11/2008         13.5         .947         6.75         7.03           11/3/2008         10.4         .855         8.06         7.68           11/24/2008         2.4         .746         12.98         8.07           12/3/2008         2.8         .633         12.12         7.94           Site         Date         Temp         Sp. Cond.         DO(mg/l)         pH           DR         3/24/2008         3.5         .438         12.10         7.86           5/5/2008         15.2         .502         8.84         7.86						
5/26/2008         19.7         .673         8.20         7.90           6/16/2008         21.8         .752         6.95         7.82           7/7/2008         22.8         .633         7.58         7.88           7/29/2008         22.1         .761         6.41         7.78           8/18/2008         20.4         .864         4.96         7.42           9/8/2008         18.2         .976         5.39         7.60           9/29/2008         18.9         .769         5.20         7.53           10/11/2008         13.5         .947         6.75         7.03           11/24/2008         2.4         .746         12.98         8.07           12/3/2008         2.8         .633         12.12         7.94           Site         Date         Temp         Sp. Cond.         DO(mg/l)         pH           DR         3/24/2008         3.5         .438         12.10         7.86           4/14/2008         7.3         .509         10.92         7.86           5/5/2008         15.2         .502         8.84         7.86           5/26/2008         17.9         .677         7.17         7.79						
6/16/2008         21.8         .752         6.95         7.82           7/7/2008         22.8         .633         7.58         7.88           7/29/2008         22.1         .761         6.41         7.78           8/18/2008         20.4         .864         4.96         7.42           9/8/2008         18.2         .976         5.39         7.60           9/29/2008         18.9         .769         5.20         7.53           10/11/2008         13.5         .947         6.75         7.03           11/3/2008         10.4         .855         8.06         7.68           11/24/2008         2.4         .746         12.98         8.07           12/3/2008         2.8         .633         12.12         7.94           Site         Date         Temp         Sp. Cond.         DO(mg/l)         pH           DR         3/24/2008         3.5         .438         12.10         7.86           4/14/2008         7.3         .509         10.92         7.86           5/5/2008         15.2         .502         8.84         7.86           5/26/2008         17.9         .677         7.17         7.79						
7/7/2008         22.8         .633         7.58         7.88           7/29/2008         22.1         .761         6.41         7.78           8/18/2008         20.4         .864         4.96         7.42           9/8/2008         18.2         .976         5.39         7.60           9/29/2008         18.9         .769         5.20         7.53           10/11/2008         13.5         .947         6.75         7.03           11/3/2008         10.4         .855         8.06         7.68           11/24/2008         2.4         .746         12.98         8.07           12/3/2008         2.8         .633         12.12         7.94           Site         Date         Temp         Sp. Cond.         DO(mg/l)         pH           DR         3/24/2008         3.5         .438         12.10         7.86           4/14/2008         7.3         .509         10.92         7.86           5/5/2008         15.2         .502         8.84         7.86           5/5/2008         17.9         .677         7.17         7.79           6/16/2008         21.8         .766         6.93         7.75						
7/29/2008       22.1       .761       6.41       7.78         8/18/2008       20.4       .864       4.96       7.42         9/8/2008       18.2       .976       5.39       7.60         9/29/2008       18.9       .769       5.20       7.53         10/11/2008       13.5       .947       6.75       7.03         11/3/2008       10.4       .855       8.06       7.68         11/24/2008       2.4       .746       12.98       8.07         12/3/2008       2.8       .633       12.12       7.94         Site Date       Temp Sp. Cond.       DO(mg/l)       pH         DR       3/24/2008       3.5       .438       12.10       7.86         4/14/2008       7.3       .509       10.92       7.86         5/5/2008       15.2       .502       8.84       7.86         5/26/2008       17.9       .677       7.17       7.79         6/16/2008       21.8       .766       6.93       7.75         7/7/2008       23.4       .606       7.44       7.88         7/29/2008       23.9       .753       6.86       7.90         8/18/2008<						
9/8/2008 18.2 .976 5.39 7.60 9/29/2008 18.9 .769 5.20 7.53 10/11/2008 13.5 .947 6.75 7.03 11/3/2008 10.4 .855 8.06 7.68 11/24/2008 2.4 .746 12.98 8.07 12/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 7.3 .509 10.92 7.86 4/14/2008 7.3 .509 10.92 7.86 5/5/2008 15.2 .502 8.84 7.86 5/26/2008 17.9 .677 7.17 7.79 6/16/2008 21.8 .766 6.93 7.75 7/7/2008 23.4 .606 7.44 7.88 7/29/2008 23.9 .753 6.86 7.90 8/18/2008 22.9 .737 6.12 7.73 9/8/2008 21.0 .747 5.27 7.86		7/29/2008			6.41	
9/29/2008 18.9 .769 5.20 7.53 10/11/2008 13.5 .947 6.75 7.03 11/3/2008 10.4 .855 8.06 7.68 11/24/2008 2.4 .746 12.98 8.07 12/3/2008 2.8 .633 12.12 7.94  Site Date Temp Sp. Cond. DO(mg/l) pH  DR 3/24/2008 3.5 .438 12.10 7.86 4/14/2008 7.3 .509 10.92 7.86 5/5/2008 15.2 .502 8.84 7.86 5/26/2008 17.9 .677 7.17 7.79 6/16/2008 21.8 .766 6.93 7.75 7/7/2008 23.4 .606 7.44 7.88 7/29/2008 23.9 .753 6.86 7.90 8/18/2008 22.9 .737 6.12 7.73 9/8/2008 21.0 .747 5.27 7.86		8/18/2008	20.4	.864	4.96	7.42
10/11/2008       13.5       .947       6.75       7.03         11/3/2008       10.4       .855       8.06       7.68         11/24/2008       2.4       .746       12.98       8.07         12/3/2008       2.8       .633       12.12       7.94         Site       Date       Temp       Sp. Cond.       DO(mg/l)       pH         DR       3/24/2008       3.5       .438       12.10       7.86         4/14/2008       7.3       .509       10.92       7.86         5/5/2008       15.2       .502       8.84       7.86         5/26/2008       17.9       .677       7.17       7.79         6/16/2008       21.8       .766       6.93       7.75         7/7/2008       23.4       .606       7.44       7.88         7/29/2008       23.9       .753       6.86       7.90         8/18/2008       22.9       .737       6.12       7.73         9/8/2008       21.0       .747       5.27       7.86		9/8/2008	18.2	.976	5.39	7.60
11/3/2008       10.4       .855       8.06       7.68         11/24/2008       2.4       .746       12.98       8.07         12/3/2008       2.8       .633       12.12       7.94         Site       Date       Temp       Sp. Cond.       DO(mg/l)       pH         DR       3/24/2008       3.5       .438       12.10       7.86         4/14/2008       7.3       .509       10.92       7.86         5/5/2008       15.2       .502       8.84       7.86         5/26/2008       17.9       .677       7.17       7.79         6/16/2008       21.8       .766       6.93       7.75         7/7/2008       23.4       .606       7.44       7.88         7/29/2008       23.9       .753       6.86       7.90         8/18/2008       22.9       .737       6.12       7.73         9/8/2008       21.0       .747       5.27       7.86		9/29/2008	18.9	.769	5.20	7.53
11/24/2008       2.4       .746       12.98       8.07         12/3/2008       2.8       .633       12.12       7.94         Site       Date       Temp       Sp. Cond.       DO(mg/l)       pH         DR       3/24/2008       3.5       .438       12.10       7.86         4/14/2008       7.3       .509       10.92       7.86         5/5/2008       15.2       .502       8.84       7.86         5/26/2008       17.9       .677       7.17       7.79         6/16/2008       21.8       .766       6.93       7.75         7/7/2008       23.4       .606       7.44       7.88         7/29/2008       23.9       .753       6.86       7.90         8/18/2008       22.9       .737       6.12       7.73         9/8/2008       21.0       .747       5.27       7.86		10/11/2008	13.5	.947	6.75	7.03
12/3/2008       2.8       .633       12.12       7.94         Site       Date       Temp       Sp. Cond.       DO(mg/l)       pH         DR       3/24/2008       3.5       .438       12.10       7.86         4/14/2008       7.3       .509       10.92       7.86         5/5/2008       15.2       .502       8.84       7.86         5/26/2008       17.9       .677       7.17       7.79         6/16/2008       21.8       .766       6.93       7.75         7/7/2008       23.4       .606       7.44       7.88         7/29/2008       23.9       .753       6.86       7.90         8/18/2008       22.9       .737       6.12       7.73         9/8/2008       21.0       .747       5.27       7.86		11/3/2008	10.4	.855	8.06	7.68
Site         Date         Temp         Sp. Cond.         DO(mg/l)         pH           DR         3/24/2008         3.5         .438         12.10         7.86           4/14/2008         7.3         .509         10.92         7.86           5/5/2008         15.2         .502         8.84         7.86           5/26/2008         17.9         .677         7.17         7.79           6/16/2008         21.8         .766         6.93         7.75           7/7/2008         23.4         .606         7.44         7.88           7/29/2008         23.9         .753         6.86         7.90           8/18/2008         22.9         .737         6.12         7.73           9/8/2008         21.0         .747         5.27         7.86				.746	12.98	8.07
DR 3/24/2008 3.5 .438 12.10 7.86 4/14/2008 7.3 .509 10.92 7.86 5/5/2008 15.2 .502 8.84 7.86 5/26/2008 17.9 .677 7.17 7.79 6/16/2008 21.8 .766 6.93 7.75 7/7/2008 23.4 .606 7.44 7.88 7/29/2008 23.9 .753 6.86 7.90 8/18/2008 22.9 .737 6.12 7.73 9/8/2008 21.0 .747 5.27 7.86		12/3/2008	2.8	.633	12.12	7.94
4/14/2008       7.3       .509       10.92       7.86         5/5/2008       15.2       .502       8.84       7.86         5/26/2008       17.9       .677       7.17       7.79         6/16/2008       21.8       .766       6.93       7.75         7/7/2008       23.4       .606       7.44       7.88         7/29/2008       23.9       .753       6.86       7.90         8/18/2008       22.9       .737       6.12       7.73         9/8/2008       21.0       .747       5.27       7.86	Site	Date	Temp	Sp. Cond.	DO(mg/l)	рН
5/5/2008       15.2       .502       8.84       7.86         5/26/2008       17.9       .677       7.17       7.79         6/16/2008       21.8       .766       6.93       7.75         7/7/2008       23.4       .606       7.44       7.88         7/29/2008       23.9       .753       6.86       7.90         8/18/2008       22.9       .737       6.12       7.73         9/8/2008       21.0       .747       5.27       7.86	DR	3/24/2008	3.5	.438	12.10	7.86
5/26/2008       17.9       .677       7.17       7.79         6/16/2008       21.8       .766       6.93       7.75         7/7/2008       23.4       .606       7.44       7.88         7/29/2008       23.9       .753       6.86       7.90         8/18/2008       22.9       .737       6.12       7.73         9/8/2008       21.0       .747       5.27       7.86		4/14/2008	7.3	.509	10.92	7.86
6/16/2008       21.8       .766       6.93       7.75         7/7/2008       23.4       .606       7.44       7.88         7/29/2008       23.9       .753       6.86       7.90         8/18/2008       22.9       .737       6.12       7.73         9/8/2008       21.0       .747       5.27       7.86		5/5/2008	15.2	.502	8.84	7.86
7/7/2008       23.4       .606       7.44       7.88         7/29/2008       23.9       .753       6.86       7.90         8/18/2008       22.9       .737       6.12       7.73         9/8/2008       21.0       .747       5.27       7.86		5/26/2008	17.9	.677	7.17	7.79
7/29/2008       23.9       .753       6.86       7.90         8/18/2008       22.9       .737       6.12       7.73         9/8/2008       21.0       .747       5.27       7.86		6/16/2008	21.8	.766	6.93	7.75
8/18/2008       22.9       .737       6.12       7.73         9/8/2008       21.0       .747       5.27       7.86		7/7/2008	23.4	.606	7.44	7.88
9/8/2008 21.0 .747 5.27 7.86		7/29/2008				
				.737	6.12	
9/29/2008 20.2 .461 6.85 7.68						
		9/29/2008	20.2	.461	6.85	7.68

10/11/2008	14.8	.510	5.10	7.27
11/3/2008	9.9	.619	5.62	7.54
11/24/2008	0.5	.765	12.82	8.06
12/3/2008	2.0	.613	12.22	7.92

Data are missing due to equipment or associated specific probes not being deployed, equipment failure, time of maintenance or calibration of equipment, or repair/replacement of a sampling station platform. Any NANs in the dataset stand for "not a number" and are the result of low power, disconnected wires, or out of range readings. If additional information on missing data is needed, contact the Research Coordinator at the reserve submitting the data.