

Tijuana River (TJR) NERR Meteorological Metadata
January – December 2023
Latest Update: 02/04/2025

I. Data Set & Research Descriptors

1) Principal investigator(s) & contact persons

Jeff Crooks, Research Coordinator

301 Caspian Way
Imperial Beach, CA 91932
Phone: (619) 575-3613
Fax: (619) 575-6913
e-mail: jcrooks@trnerr.org

Justin McCullough, Research Associate

301 Caspian Way
Imperial Beach, CA 91932
Phone: (619) 575-3613 x321
e-mail: jmccullough@trnerr.org

2) Entry verification

Data are uploaded from the CR1000 data logger to a personal computer with a Windows 7 or newer operating system. Files are exported from LoggerNet in a comma-delimited format and uploaded to the CDMO where they undergo automated primary QAQC and become part of the CDMO's online provisional database. During primary QAQC, data are flagged if they are missing or out of sensor range. The edited file is then returned to the reserve where it is opened in Microsoft Excel and processed using the CDMO's NERRQAQC Excel macro. The macro inserts station codes, creates metadata worksheets for flagged data and summary statistics, and graphs the data for review. It allows the user to apply QAQC flags and codes to the data, append files, and export the resulting data file to the CDMO for tertiary QAQC and assimilation into the CDMO's authoritative online database. For more information on QAQC flags and QAQC codes, see Sections 11 and 12. Justin McCullough is responsible for all data management.

3) Research objectives:

The Tijuana River National Estuarine Research Reserve (TRNERR) represents the largest, most intact coastal marsh system remaining in Southern California. It has contiguous beach, dune, tidal channel, mudflat, marsh, transitional, and upland habitat. It is also home to numerous threatened and endangered species. Because of its highly urbanized setting, situated between the cities of Tijuana, Baja California, Mexico, and San Diego, California, USA, it is heavily impacted. A primary management issue is transboundary flows of the Tijuana River, which convey anthropogenic pollutants (primarily associated with partially-treated and untreated wastewater), nutrients, and sediment. The principal objective is to record long-term and episodic meteorological data for the Tijuana River Estuary in order to observe any environmental changes or trends over time. Data are also used as corollary information in ongoing biological, hydrological and geographical studies being conducted at the reserve. Of particular management interest is rainfall, which can trigger large flow events in the Tijuana River.

4) Research Methods

A model CR1000 Campbell Scientific datalogger samples meteorological sensors every 5 seconds. At 15 minute intervals, averages, totals, or instantaneous readings are taken, depending upon sensor type, and written to a storage table on the CR1000. Data are reported in Pacific Standard Time (PST).

Monthly, sensors on the weather station are inspected for damage or debris. If any is found, it is repaired and/or cleaned. Sensors are removed and sent back to Campbell Scientific for calibration at minimum of every one to two years, depending on the sensor.

Campbell Scientific data telemetry equipment was installed at the Tidal Linkage station on 06/12/2006 and transmits data to the NOAA GOES satellite, NESDIS ID #3B01468A. The transmissions are scheduled hourly at 0:47:00 after the hour and contain four (4) data sets reflecting fifteen minute data sampling intervals. Upon receipt by the CDMO, the data undergoes the same automated primary QA/QC process detailed in Section 2 above. The "real-time" telemetry data become part of the provisional dataset until undergoing secondary and tertiary QA/QC and assimilation in the CDMO's authoritative online database. Provisional and authoritative data are available at <http://cdmo.baruch.sc.edu>.

The 15 minute Data are collected in the following formats for the CR1000:

- Averages from 5-second data:
 - Air Temperature (°C), Relative Humidity (%), Barometric Pressure (mb), Wind Speed (m/s), Wind Direction (degrees), and Battery Voltage (volts)
- Maximum and Minimum Air Temperature (°C) and times from 5-second data (these data are available from the reserve)
- Maximum Wind Speed (m/s) and time from 5-second data
- Wind Direction Standard Deviation (degrees) from 5-second data
- Totals:
 - Precipitation (mm), PAR (millimoles/m²), and Cumulative Precipitation (mm; cumulative precipitation is no longer available via export from the CDMO. Please contact the Reserve or the CDMO for more information or to obtain these data.)

Recommended calibration frequency for the MET station sensors:

- Temperature/Humidity - yearly recalibration
- Rain Gauge - yearly recalibration
- Wind Speed/Direction - every 2 years factory maintenance
- Barometric Pressure - every 2 years recalibration
- Photosynthetically Active Radiation (PAR) - Apogee Quantum Sensor - every 2 years recalibration
- CR1000 - every 5 years

5) Site location and character:

The Tijuana River NERR is located on the southern Pacific coast, next to the California border with Mexico at a latitude of 32°34'N and longitude of 117°07'W. The area surrounding the 2,531 acre reserve is heavily developed by residential housing, as is the watershed which drains into the estuary. Approximately two-thirds of the watershed resides in Mexico and is subject to periodic raw sewage outflows. The northeastern section is bordered by a military helicopter training base. Vegetation in the area is dominated by common pickleweed (*Salicornia pacifica*) and Pacific cordgrass (*Spartina foliosa*).

The weather station is located approximately 30m west of the TRNERR Visitor Center at a latitude of 32°34'28.5"N and longitude of 117°07'37.3"W. The station is about 800m northeast of the water quality sampling station at Oneonta Slough. The vegetation surrounding the weather station is mainly upland scrub species.

The anemometer (2.8m), wind direction (2.8m), and PAR (2.7m) sensors are attached to a 3.5m aluminum tower. The temperature and humidity sensors are

located midway up (2.2m) and on the west side of the tower. The barometric pressure sensor is mounted in the CR1000 enclosure at a height of 1.5m. The Tipping Bucket rain gauge (relocated in 2014) is attached to the fence 1.7 meters to the northwest of the main tower. It is 2.2m above the ground to limit interference from the security fence surrounding the weather station. The sensors are wired to the CR1000 following the protocol in the CDMO MET SOP.

SWMP Station Timeline and Details:

Station Code	SWMP Status	Station Name	Location	Active Dates	Reason Decommissioned	Notes
TL	Primary	Tidal Linkage	32°34'28.5"N 117°07'37.3"W	Jan 2001 - present	N/A	None
Tower/sensors installed		Height above ground (m)	Notes			
Tower		3.5				
Temperature/Relative Humidity		2.2				
Barometric Pressure		1.5	sensor resides within enclosure			
Wind		2.8				
PAR		2.7				
Precipitation gauge		2.2	1.7m northwest of tower			

6) Data collection Period

Data was collected for all parameters at the station from 01/01/2023 00:00 and continued through 12/31/2023 23:45.

File Start Date and Time	File End Date and Time
12/15/2022 13:00	1/19/2023 14:15
1/19/2023 14:30	2/16/2023 13:30
2/16/2023 13:45	3/16/2023 13:15
3/16/2023 13:30	4/14/2023 13:15
4/14/2023 13:30	5/15/2023 13:15
5/15/2023 13:30	6/15/2023 9:30
6/15/2023 9:45	7/14/2023 11:30
7/14/2023 11:45	8/14/2023 10:00
8/14/2023 10:15	9/14/2023 9:00
9/14/2023 9:15	10/16/2023 13:30
10/16/2023 13:45	11/16/2023 13:15
11/16/2023 13:30	12/14/2023 11:15
12/14/2023 11:30	1/12/2024 12:30

7) Distribution

NOAA retains the right to analyze, synthesize and publish summaries of the NERRS System-wide Monitoring Program data. The NERRS retains the right to be fully credited for having collected and processed the data. Following academic courtesy standards, the NERR site where the data were collected should be contacted and fully acknowledged in any subsequent publications in which any part of the data are used. The data set enclosed within this

package/transmission is only as good as the quality assurance and quality control procedures outlined by the enclosed metadata reporting statement. The user bears all responsibility for its subsequent use/misuse in any further analyses or comparisons. The Federal government does not assume liability to the Recipient or third persons, nor will the Federal government reimburse or indemnify the Recipient for its liability due to any losses resulting in any way from the use of this data.

Requested citation format:

NOAA National Estuarine Research Reserve System (NERRS). System-wide Monitoring Program. Data accessed from the NOAA NERRS Centralized Data Management Office website: <http://www.nerrsdata.org/>; accessed 12 October 2023

NERR meteorological data and metadata can be obtained from the Research Coordinator at the individual NERR site (please see Principal Investigators and Contact Persons), from the Data Manager at the Centralized Data Management Office (please see personnel directory under the general information link on the CDMO home page) and online at the CDMO home page www.nerrsdata.org. Data are available in comma delimited format.

8) Associated researchers and projects

The meteorological station is part of the NERRS System Wide Monitoring Program (SWMP) that also includes data collection for water quality using YSI data loggers for in situ measurements (temperature, salinity, dissolved oxygen, turbidity, pH and water depth) and collecting monthly grab and diel water samples to analyze nutrient/pigment contents (orthophosphate, ammonium, nitrate/nitrite, chlorophyll) in the laboratory. These data may be correlated with this meteorological dataset and are available at www.nerrsdata.org.

II. Physical Structure Descriptors

9) Sensor specifications

Photosynthetically Active Radiation (PAR) Apogee Quantum Sensor

Units: mmoles/m^2 (total flux)

Sensor type: High stability silicon photovoltaic detector (blue enhanced)

Model#: SQ-110

Light spectrum waveband: 410 to 655 nm

Temperature dependence: $.06 \pm .06\%$ per $^{\circ}\text{C}$

Stability: $\pm 2\%$ change over 1 yr

Operating Temperature: -40°C to 70°C ; Humidity: 0 to 100%

Sensitivity: $0.2 \text{ mV per } \mu\text{mol m}^{-2} \text{ s}^{-1}$

Multiplier: 0.025

Serial#: 18025

Date of last calibration: March 2021

Dates in service: 04/20/2021 - 04/14/2023

Serial#: 12144

Date of last calibration: 03/21/2023

Dates in service: 04/14/2023 - current as of 12/31/2023

Wind speed

Units: meter per second (m/s)

Sensor type: 18 cm diameter 4-blade helicoids propeller molded of polypropylene

Model#: R.M. Young 05106 Wind Monitor - Marine

Range: 0-100 m/s (224 mph)

Accuracy: $\pm 1\%$

Serial#: WM 150921

Date of last factory calibration: November 11, 2020

Dates in service: 12/15/2020 - 03/16/2023

Wind direction

Units: degrees

Sensor type: balanced vane, 38 cm turning radius

Model#: R.M. Young 05106 Wind Monitor - Marine

Range: 360° mechanical, 355° electrical (5° open)

Accuracy: $\pm 3^\circ$

Serial#: WM 150921

Date of last factory calibration: November 11, 2020

Dates in service: 12/15/2020 - 03/16/2023

Wind speed

Units: meter per second (m/s)

Sensor type: 18 cm diameter 4-blade helicoids propeller molded of polypropylene

Model#: R.M. Young 05103 Wind Monitor

Range: 0-100 m/s (224 mph)

Accuracy: $\pm 1\%$

Serial#: 57955

Date of last factory calibration: January 4, 2023

Dates in service: 03/16/2023 - current as of 12/31/2023

Wind direction

Units: degrees

Sensor type: balanced vane, 38 cm turning radius

Model#: R.M. Young 05103 Wind Monitor

Range: 360° mechanical, 355° electrical (5° open)

Accuracy: $\pm 3^\circ$

Serial#: 57955

Date of last factory calibration: January 4, 2023

Dates in service: 03/16/2023 - current as of 12/31/2023

Temperature and Relative Humidity

Units: degrees Celsius ($^\circ\text{C}$) and percent (%)

Model#: EE181-M1A1

Operating Temperature: -40 to $+60^\circ\text{C}$

Temperature Measurement Range: -40 to $+60^\circ\text{C}$

Temperature Accuracy: $\pm 0.2^\circ\text{C}$ (@ 23°C)

Relative Humidity Measurement Range: 0-100% non-condensing

RH Accuracy:

$\pm (1.3 + 0.003 \cdot \text{RH reading}) \% \text{ RH}$ (at -15° to $+40^\circ\text{C}$, 0 to 90% RH)

$\pm 2.3\% \text{ RH}$ (at -15° to $+40^\circ\text{C}$, 90 to 100% RH)

$\pm (1.4 + 0.01 \cdot \text{RH reading}) \% \text{ RH}$ (at -25° to $+60^\circ\text{C}$)

$\pm (1.5 + 0.015 \cdot \text{RH reading}) \% \text{ RH}$ (at -40° to $+60^\circ\text{C}$)

Note: This sensor caps relative humidity values at 100%, measured values

>100% are altered to 100%

Serial#: 21091600204263
Date of last calibration: 11/03/2021
Dates in service: 05/17/2022 - 05/15/2023

Serial#: 210916002058D9
Date of last calibration: 04/14/2023
Dates in service: 05/15/2023 - current as of 12/31/2023

Barometric Pressure

Units: millibars (mbar)
Model#: PTB110 (CS 106)
Operating Temperature: -40 to +60C
Pressure Measurement Range: 500-1100 mb
Humidity: non-condensing
Accuracy: ± 0.3 mb at +20 °C
Stability: ± 0.1 mb per year

Serial#: L3220432
Date of Last calibration: 02/14/2022
Dates in service: 05/18/2022 - 11/17/2023

Serial#: G1760044
Date of Last calibration: 09/01/2022
Dates in service: 11/17/2023 - current as of 12/31/2023

Precipitation

Units: millimeters (mm)
Sensor type: Tipping Bucket Rain Gauge
Model#: TR-525-U-01
Rainfall per tip: 0.01 inch
Operating range: Temperature: 0 to 70°C; Humidity: 0 to 100%
Accuracy: $\pm 1.0\%$ up to 2 in/hr

Serial#: 89520-0122
Dates of calibration: February 7, 2022; August 30, 2022; August 24, 2023
Dates in service: 02/07/2022- current as of 12/31/2023

Datalogger

Model: CR1000
Specs: The CR1000 has a 2MB Flash EEPROM that is used to store the Operating System. Another 128K Flash is used to store configuration settings. A minimum of 2MB SRAM (4MB optional) is available for program storage (16K), operating system use, and data storage. Additional storage is available by using a compact flash card in the optional CFM100 Compact Flash Module.
CR1000 firmware/OS version: cr1000.Std.32.04; updated when calibrated on 12/13/2019
CR1000 Program Versions: tjr1met_CR1000_6.2_022015.cr1
Serial#: 5478
Date of Calibration: 12/13/2019
Dates in service: 01/10/2020 - current as of 12/31/2023

GOES Transmitter

Campbell Scientific, Inc. HDR GOES Transmitter
Model Number: TX312
Serial Number: 1351
Date Installed: data of installation unknown

10) Coded variable definitions

Sampling station:	Sampling site code:	Station code:
Tidal Linkage	TL	tjrtlmet

11) QAQC flag definitions

QAQC flags provide documentation of the data and are applied to individual data points by insertion into the parameter's associated flag column (header preceded by an F_). During primary automated QAQC (performed by the CDMO), -5, -4, and -2 flags are applied automatically to indicate data that is above or below sensor range, or missing. All remaining data are then flagged 0, as passing initial QAQC checks. During secondary and tertiary QAQC, 1, -3, and 5 flags may be used to note data as suspect, rejected due to QAQC, or corrected, respectively.

- 5 Outside High Sensor Range
- 4 Outside Low Sensor Range
- 3 Data Rejected due to QAQC
- 2 Missing Data
- 1 Optional SWMP supported parameter
- 0 Passed Initial QAQC Checks
- 1 Suspect Data
- 2 Open - reserved for later flag
- 3 Open - reserved for later flag
- 4 Historical Data: Pre-Auto QAQC
- 5 Corrected Data

12) QAQC code definitions

QAQC codes are used in conjunction with QAQC flags to provide further documentation of the data and are also applied by insertion into the associated flag column. There are three (3) different code categories, general, sensor, and comment. General errors document general problems with the CR1000, sensor errors are sensor specific, and comment codes are used to further document conditions or a problem with the data. Only one general or sensor error and one comment code can be applied to a particular data point, but some comment codes (marked with an * below) can be applied to the entire record in the F_Record column.

General Errors

GIM	Instrument malfunction
GIT	Instrument recording error, recovered telemetry data
GMC	No instrument deployed due to maintenance/calibration
GMT	Instrument maintenance
GPD	Power down
GPF	Power failure / Low battery
GPR	Program reload
GQR	Data rejected due to QA/QC checks
GSM	See metadata

Sensor Errors

SDG	Suspect due to sensor diagnostics
SIC	Incorrect calibration constant, multiplier or offset
SIW	Incorrect wiring
SMT	Sensor maintenance
SNV	Negative value
SOC	Out of calibration
SQR	Data rejected due to QAQC checks

SSD Sensor drift
SSN Not a number / unknown value
SSM Sensor malfunction
SSR Sensor removed

Comments

CAF Acceptable calibration/accuracy error of sensor
CCU Cause unknown
CDF Data appear to fit conditions
CML Snow melt from previous snowfall event
CRE* Significant rain event
CSM* See metadata
CVT* Possible vandalism/tampering
CWE* Significant weather event

13) Other remarks / notes

Data are missing due to equipment or associated specific sensors not being deployed, equipment failure, time of maintenance or calibration of equipment, or repair/replacement of a sampling station platform. Any NaNs in the dataset stand for "not a number" and are the result of low power, disconnected wires, or out of range readings. If additional information on missing data is needed, contact the Research Coordinator at the reserve submitting the data.

Relative Humidity data greater than 100 are within range of the sensor accuracy of +/-3% and are flagged and coded as suspect, <1> (CAF). Values greater than 103 are rejected <-3>.

Data recorded for all parameters (with the exception of cumulative precipitation) at the midnight timestamp (00:00) are the 15 minute averages and totals for the 23:45-23:59 time period of the previous day. Cumulative precipitation data at the midnight timestamp (00:00) are the sum of raw (unrounded) precipitation data from 00:00 to 23:59 of the previous day. Summing each individual 15-minute total precipitation value from the same period will result in small differences from cumulative precipitation due to rounding. It is especially important to note how data at the midnight timestamp are recorded when using January 1st and December 31st data. **Note: Cumulative precipitation is no longer available via export from the CDMO. Please contact the Reserve or the CDMO for more information or to obtain these data.**

All periods where sensor values deviated noticeably from the rest of the data set, e.g. periods of high or low temperature, low barometric pressure etc., were evaluated against data from nearby weather stations and also compared to local daily analog measurements when available. All such periods were found to be comparable with the ancillary data, unless noted otherwise, below.

Additional Flagging Notes:

It was discovered in early January 2023 that the wiring for the tipping bucket rain gauge was damaged, resulting in several rain events not being recorded. These rain events were captured at nearby weather stations (www.wunderground.com) and are presented below. **TotPrcp** and **CumPrcp** from **12/15/2022 13:00** thru **1/12/2023 00:00** were rejected <-3>[SSM] (CSM).

KCAIMPER28		KCAIMPER24
0.16	12/27/22	0.10
0.69	12/28/22	0.57
0.02	12/31/22	0.05
1.01	1/1/23	1.12
0.06	1/2/23	0.02
0.10	1/3/23	0.13
0.12	1/5/23	0.22
0.01	1/6/23	0.02
0.38	1/10/23	0.25
0.01	1/11/23	0.01

On **1/19/2023** between **12:15** and **14:15**, the **wind** set was replaced with a newly calibrated set. After the installation, the spinning of the propeller was audible, so the old wind set was reinstalled. This maintenance resulted in the rejection **<-3>[SMT] (CSM)** of **WSpd, Wdir, SDWDir, and MaxWSpd** for the above time frame.

On **3/16/2023** between 12:35 and 13:03, the **wind** set was replaced with a newly calibrated set. This maintenance resulted in the rejection **<-3> [SMT] (CSM)** of **WSpd, Wdir, SDWDir, and MaxWSpd** from **12:45** to **13:15**.

During cleaning and maintenance of the tipping bucket rain gauge on **3/16/2023**, the gauge was activated, resulting in erroneous rain data at **12:45**. **TotPrcp** was corrected **<5>[SMT] (CSM)** at this time. **CumPrcp** was corrected and flagged **<5>[SMT] (CSM)** from **3/16/2023 12:45** to **3/17/2023 00:00**.

The **PAR** sensor was replaced with a newly calibrated sensor on **4/14/2023** between 12:25 and 12:50, resulting in data being rejected **<-3>[SMT] (CSM)** from **12:30** to **13:00**.

The **temp/rh** probe was swapped for a newly calibrated probe between 13:08 and 13:14, resulting in **5/15/2020 13:15** being rejected **<-3>[SMT] (CSM)**.

The wind set was cleaned, resulting in **WSpd, Wdir, SDWDir, and MaxWSpd** being rejected **<-3>[SMT] (CSM)** at **6/15/2023 9:30**.

Only upon closer examination of the meteorological data while conducting the yearly QAQC, was it noticed that the **barometric pressure (BP)** continually drifted to lower values when compared to other nearby sites. This occurred during three separate intervals beginning at approximately **8/27/2023 12:45**. During the following month's maintenance at **9/14/2023 9:00**, the values suddenly returned to comparable values. The data show the same general trends of increasing and decreasing barometric pressure as seen at the nearby weather station, but continually show a downward trend. This again occurred upon servicing the equipment at **10/12/2023 11:15**, to only seemingly be fixed upon another visit to the equipment at **10/16/2023 13:45**, at which point the sensor appears to return to functioning properly. Yet again in late October, the downward trend was observed, resulting in rejected data. All of this data with the downward trend is rejected **<-3>[SSM] (CSM)**. Data that appears to closely match our nearby weather station located at Los Peñasquitos Lagoon during this time period is marked suspect **<1>[SSM] (CSM)**. The BP sensor was swapped with a new one at **11/17/2023 14:10**, resulting in data being rejected **<-3>[SMT] (CSM)** at

11/17/2023 14:15. The following are the time spans during which data were flagged as rejected and suspect.

Rejected data:

08/27/2023 12:45 - 09/14/2023 8:45
10/12/2023 11:15 - 10/16/2023 13:45
10/26/2023 00:00 - 11/16/2023 13:00

Suspect data:

09/14/2023 09:00 - 10/12/2023 11:00
10/16/2023 14:00 - 10/25/2023 23:45
11/16/2023 13:15 - 11/17/2023 14:00