

Tijuana River (TJR) NERR Water Quality Metadata
January to December 2023
Latest Update: April 14, 2024

Note: This is a provisional metadata document; it has not been authenticated as of its download date. Contents of this document are subject to change throughout the QAQC process and it should not be considered a final record of data documentation until that process is complete. Contact the CDMO (cdmosupport@baruch.sc.edu) or reserve with any additional questions.

I. Data Set and Research Descriptors

1) Principal investigator(s) and contact persons –

Jeff Crooks, Research Coordinator

301 Caspian Way
Imperial Beach, CA 91932
Phone: (619) 575-3613
Fax: (619) 575-6913
E-mail: jcrooks@tnerr.org

Monica Almeida, Research Associate

301 Caspian Way
Imperial Beach, CA 91932
Phone: (619) 575-3613
Fax: (619) 575-6913
E-mail: malmeida@tnerr.org

Justin McCullough, Research Associate

301 Caspian Way
Imperial Beach, CA 91932
Phone: (619) 575-3613
Fax: (619) 575-6913
E-mail: jmccullough@tnerr.org

2) Entry verification –

Deployment data are uploaded from the YSI data logger to a personal computer with Windows 7 or newer operating system. Files are exported from EcoWatch in a comma-delimited format (.CDF), EcoWatch Lite in a comma separated file (CSV) or KOR Software in a comma separated file (CSV) and uploaded to the CDMO where they undergo automated primary QAQC; automated Depth/Level corrections for changes in barometric pressure (cDepth or cLevel parameters); and become part of the CDMO's online provisional database. All pre- and post-deployment data are removed from the file prior to upload. During primary QAQC, data are flagged if they are missing or out of sensor range. The edited file is then returned to the reserve for secondary QAQC where it is opened in Microsoft Excel and processed using the CDMO's NERRQAQC Excel macro. The macro inserts station codes, creates metadata worksheets for flagged data and summary statistics, and graphs the data for review. It allows the user to apply QAQC flags and codes to the data, remove any overlapping deployment data, append files, and export the resulting data file for upload to the CDMO. Upload after secondary QAQC results in ingestion into the database as provisional plus data, recalculation of cDepth or cLevel parameters, and finally tertiary QAQC by the CDMO and assimilation into the CDMO's authoritative online database. Where deployment overlap occurs between

files, the data produced by the newly calibrated sonde is generally accepted as being the most accurate. For more information on QAQC flags and codes, see Sections 11 and 12. Monica Almeida is responsible for data management.

2) Research objectives –

The Tijuana River National Estuarine Research Reserve (TRNERR) represents the largest, most intact coastal marsh system remaining in Southern California. It has contiguous beach, dune, tidal channel, mudflat, marsh, transitional, and upland habitat. It is also home to numerous threatened and endangered species. TRNERR experiences a Mediterranean climate, with warm, dry summers and cool, wet winters. The majority of rainfall occurs between the months of October and March. Because of its highly urbanized setting, situated between the cities of Tijuana, Baja California, Mexico, and San Diego, California, USA, it is heavily impacted. A primary management concern is transboundary flows of the Tijuana River, which convey anthropogenic pollutants (primarily associated with partially-treated and untreated wastewater), nutrients, and sediment. About a quarter of the reserve's 2,531 acres are tidally influenced and few channels are deep enough for datalogger deployment. Currently, there are two SWMP stations located within the TRNERR boundaries, and two SWMP stations are located nearby in south San Diego Bay. Station locations are designed to investigate spatial patterns of water quality parameters, with comparisons between the Tijuana Estuary and San Diego Bay. In addition, telemetry of Tijuana Estuary water quality stations informs management action, particularly related to potential closure of the tidal inlet by wave-driven accumulation of sediment. Mouth closures (detected by cessation of tidal action as indicated with the water level sensors on the dataloggers) can cause anoxia, mortality of fish and shellfish, and flooding.

Two stations were originally set up: a “control” station Oneonta Slough (OS), which is still in place, was established on the northern end of Oneonta Slough, relatively far away from the main source of river-borne pollution. Another station, River Channel (RC), was situated in a site most affected by sewage outflow. Datalogger deployment at RC, however, was continually interrupted by both shifting sediment and massive wracks of kelp (*Macrocystis pyrifera*), which would often bury the deployment set-up on incoming tides. After a number of different deployment designs were implemented without success, data collection at the RC site was terminated in 2004. Another station was located at the inlet to the Model Marsh (MM), a constructed 20-acre restoration site in the southern arm of the estuary. The Model Marsh was opened to tidal flushing in February 2000 and datalogging at the station began in October 2000. The site was discontinued in January 2008, again due to heavy sedimentation. The Boca Rio (BR) site is located near the mouth of the Tijuana River, although the mouth has migrated south in recent years. The BR station was established in December 2004 to replace the RC station, and remains active and is the site closest to the interface of the river and ocean.

The South Bay (SB) station was established in January 2008 and is located at the mouth of Otay River, which flows into South San Diego Bay. It is within the San Diego National Wildlife Refuge Complex, which also includes portions of the TRNERR. It was established to document conditions in the south bay, particularly associated with a marsh restoration in the adjacent salt ponds. The Pond Eleven (PE) station was located in a non-tidal salt pond adjacent to the South Bay logger. A tide gate was the only source of water into the pond, which was one of the first in a series of ponds with increasing salinities. The PE sonde was deployed from July 2008 to September 2010. The US Fish and Wildlife Service began restoration of this area, including Pond Eleven, from September 2010 to its completion in October 2011. A levee was breached to open Pond Eleven to the bay, which made the area tidal, and channels were excavated to further enhance circulation. Due to extensive restoration, the site had to be relocated. Datasonde deployments and nutrient data collection began in January 2012 at a new location site named Pond Restored (PR) (the name change was warranted because of the different location and profound differences in the pond before and after restoration). The Pond Restored datalogger is located approximately 560 meters southwest from where the Pond Eleven datalogger was originally. The images below show pre- and post-restoration of the salt ponds and the station locations. The post restoration photo includes the PE site as a reference to the new PR site. No sampling currently occurs at the PE site.



3) Research methods –

Each datalogger deployment site is comprised of a 4-inch diameter PVC/ABS pipe strapped vertically to either two “rail” style fence posts or a 2-inch diameter galvanized steel pole (depending on the channel substrate - i.e. sand/cobble or silty/clayey sediment) that have been driven into the sediment. Multiple 1.5-inch diameter holes have been drilled around the bottom length of the PVC pipe to permit unrestricted water flow to the sensors. Upon deployment, the datalogger is placed into and rests on a bolt fixed across the bottom of the pipe.

The sampling period is between four to six weeks, with measurements taken every 15 minutes. Measurements for specific conductivity, salinity, dissolved oxygen (percent saturation), dissolved oxygen (mg/l), temperature, turbidity, pH and water level are recorded. Prior to deployment, each sensor is calibrated with its associated calibration standard(s)/method in the laboratory. At the end of each sampling period, the YSI dataloggers are brought back to the laboratory for data downloading, cleaning and recalibration. These procedures are carried out according to the methods described in the YSI Operations Manual (see sections 3 and 7). Calibration standards for specific conductivity (50mS/cm) and turbidity (124 FNU) are purchased from YSI, and pH standards (7 and 10) are manufactured by Ricca Chemical Company and purchased from Fisher Scientific. On January 18, we started to report Chlorophyll data collected at the Oneonta Slough site. The Total Algae sensor goes through a two-point calibration, using DI water and Rhodamine WT dye, which is purchased from Kingscote Chemicals. The QA/QC procedures for the collected data are followed from the CDMO Operations Manual version 6.6 – February 2015.

In the field, concurrent with the datalogger’s deployment, the YSI ProDSS multiparameter water quality meter is used to collect data for comparison. Parameters such as specific conductivity, salinity, DO (percent saturation and mg/l), temperature and barometric pressure, are measured and recorded. Once a month, the handheld meter is calibrated in the specific conductivity standard (50mS/cm) and 100% oxygen saturated water before each use.

In 2015, TRNERR started to report level data. In January and February, datalogger holders were surveyed using the Spectra Precision Epoch50 real-time kinematic (RTK) GPS and calculations were done to find the correct depth offset. In December of 2017, Boca Rio and Oneonta Slough datasondes were resurveyed after cleaning the logger holder and switching from YSI 6-series datasondes to YSI EXO2. In these surveys, RTK GPS data used the NAD_1983_2011_StatePlane_California_VI_FIPS_0406 (meters) projected coordinate system, NAVD88 vertical datum (meters), and GEOID 12A model.

In late November and early December of 2021, all stations were re-surveyed with Spectra Precision SP80 RTK GPS. Surveys were set to NAD_1983_2011_StatePlane_California_VI_FIPS_0406 (meters) projected coordinate system, NAVD88 vertical datum (meters), and GEOID 18 model. The sondes' offsets over the years are documented in the "Specific Site Characteristics" section.

Site Name	Boca Rio
Site infrastructure description	Datalogger is deployed in a PVC/ABS holder strapped vertically to two "rail" style fence posts driven into a sand (with some silt and clay) substrate
Surveying equipment	Spectra Precision Epoch50 real-time kinematic (RTK) GPS
Survey monument used	TJR NERR local benchmarks installed by a contractor company and leveled to the NGS reference benchmark U 1305 (PID - DC1330)
Survey occupation date	January 27, 2015
Survey occupation duration	Approx. 10 minutes
Ellipsoid height	-35.028 (from top of battery cap)
"Quick Check" marker for deployment tube	Mark on the pole
"Quick Check" for sonde being deployed at the same location	Sonde rests on a bolt
Annual resurveying	November 15, 2021 -0.065m NAVD88 -35.164 Ellipsoid height (from top of battery cap) Spectra Precision SP80 real-time kinematic (RTK) GPS

Site Name	Oneonta Slough
Site infrastructure description	Datalogger is deployed in a PVC/ABS holder strapped vertically to a 2-inch diameter galvanized steel pole driven into a silty clay substrate
Surveying equipment	Spectra Precision Epoch50 real-time kinematic (RTK) GPS
Survey monument used	TJR NERR local benchmarks installed by a contractor company and leveled to the NGS reference benchmark U 1305 (PID - DC1330)
Survey occupation date	January 23, 2015
Survey occupation duration	Approx. 10 minutes
Ellipsoid height	-34.677 (from top of battery cap)
“Quick Check” marker for deployment tube	Mark on the pole
“Quick Check” for sonde being deployed at the same location	Sonde rests on a bolt
Annual resurveying	November 2, 2021 0.263m NAVD88 -34.544 Ellipsoid height (from top of holder) Spectra Precision SP80 real-time kinematic (RTK) GPS

Site Name	Pond Restored
Site infrastructure description	Datalogger is deployed in a PVC/ABS holder strapped vertically to a 2-inch diameter galvanized steel pole driven into a silty clay substrate
Surveying equipment	Spectra Precision Epoch50 real-time kinematic (RTK) GPS
Survey monument used	Local benchmarks installed by a contractor company and leveled to the NGS reference benchmark U 1305 (PID - DC1330)
Survey occupation date	February 25, 2015
Survey occupation duration	Approx. 10 minutes
Ellipsoid height	-34.678 (from top of holder)
“Quick Check” marker for deployment tube	Mark on the pole
“Quick Check” for sonde being deployed at the same location	Sonde rests on a bolt
Annual resurveying	November 30, 2021 -0.397m NAVD88 -34.763 Ellipsoid height (from top of holder) Spectra Precision SP80 real-time kinematic (RTK) GPS

Site Name	South Bay
Site infrastructure description	Datalogger is deployed in a PVC/ABS holder strapped vertically to a 2-inch diameter galvanized steel pole driven into a silt and clay substrate
Surveying equipment	Spectra Precision Epoch50 real-time kinematic (RTK) GPS
Survey monument used	Local benchmarks installed by a contractor company and leveled to the NGS reference benchmark U 1305 (PID - DC1330)
Survey occupation date	February 25, 2015
Survey occupation duration	Approx. 10 minutes
Ellipsoid height	-35.276 (from top of holder)
“Quick Check” marker for deployment tube	Mark on the pole
“Quick Check” for sonde being deployed at the same location	Sonde rests on a bolt
Annual resurveying	December 1, 2021 -0.629m NAVD88 -35.450 Ellipsoid height (from top of battery cap) Spectra Precision SP80 real-time kinematic (RTK) GPS

A Sutron Sat-Link2 transmitter was installed at the Oneonta Slough station on 12/20/2006 and was replaced by a newer YSI WaterLog Storm 3 transmitter on 06/08/2021. The station data is transmitted to the NOAA GOES satellite, NESDIS ID #3B0252F2. A YSI WaterLog Storm 3 transmitter was installed at the South Bay station on 10/11/2023, and the data is transmitted to the NOAA GOES satellite, NESDIS ID # 3B05762A. The transmissions are scheduled hourly and contain four (4) data sets reflecting fifteen-minute data sampling intervals. Upon receipt by the CDMO, the data undergoes the same automated primary QAQC process detailed in Section 2 above. The “real-time” telemetry data become part of the provisional dataset until undergoing secondary and tertiary QAQC and assimilation into the CDMO’s authoritative online database. Provisional and authoritative data are available at <http://cdmo.baruch.sc.edu>.

5) Site location and character –

General site Characteristics (TRNERR)

- Latitude and Longitude: 32° 34' N, 117° 7' W
- Tidal range: approximately -2ft to + 7ft MLLW
- Water depth: approximately 0.2m to 2m
- Type and amount of freshwater input: the dominant freshwater source to the estuary is the Tijuana River, which drains a 4,483km² watershed, approximately 2/3 of which resides in Mexico. Stream flows in the river vary considerably from season to season and year to year, with no flow during many months and a mean annual discharge of 0.82m³/s. Additional freshwater sources are storm drains located mostly in the northern arm of the estuary from the adjacent military airfield and residential area. The entire estuary is shallow and has a relatively small tidal prism (0.36 Mm³), so even low freshwater flows result in reduced salinity throughout the reserve. Estimated

residence times for freshwater entering the estuary vary from 7 hours to a few days, depending on the tide and mouth conditions. Rainfall within the watershed accounts for most of the freshwater entering the reserve, with 90% of the mean annual rainfall occurring between November and April. Vegetation in the area is dominated by common pickleweed (*Salicornia pacifica*) and Pacific cordgrass (*Spartina foliosa*).

e) Pollutants in area: Freshwater discharge with untreated sewage occurs year-round. Although these have decreased with the construction of a binational water treatment plant, repeated infrastructure failures have increased the frequency of raw sewage discharged throughout the year affecting the water quality.

Specific Site characteristics: Boca Rio (BR)

a) Location of site: the datalogger station is located approximately 400m north of the Tijuana River in the middle of a channel which runs north-south; 32° 33' 33.7" N, 117° 7' 44.3" W.

b) Elevation of sonde's depth port:

January 27, 2015 - December 12, 2017: 0.053m NAVD88, approximately 0.5m above the channel bottom.

December 12, 2017 - November 15, 2021: -0.056m NAVD88, approximately 0.5m above the channel bottom.

November 15, 2021 – present: -0.065m NAVD88, approximately 0.5m above the channel bottom.

c) Water depth: approximately 0.5m to 2.7m.

d) Channel width: approximately 30 m.

e) Bottom type: sand, some silt and clay.

f) Salinity: 1 ppt (extreme rain events) to 36 ppt (average of 33ppt)

Specific Site characteristics: Oneonta Slough (OS)

a) Location of site: the datalogger station is located on the upper portion of the Oneonta Slough in the northwest corner of the reserve, approximately 1.4km north of the Tijuana River. OS station is established in the middle of the same channel as the Boca Rio site; 32° 34' 6.0" N, 117° 7' 52.6" W.

b) Elevation of sonde's depth port:

January 23, 2015 - December 5, 2017: 0.332m NAVD88, approximately 0.5m above the channel bottom.

December 5, 2017 - November 2, 2021: 0.295m NAVD88, approximately 0.5m above the channel bottom.

November 2, 2021 – present: 0.263m NAVD88, approximately 0.5m above the channel bottom.

c) Water depth: approximately 0.7m to 2.4m.

d) Channel width: approximately 23m.

e) Bottom type: silty clay.

f) Salinity: 1 ppt (extreme rain events) to 39 ppt (average of 32 ppt)

g) The area adjacent to the west side of the channel is developed. There is a 50+ meter buffer of natural vegetation between development and the channel. The area adjacent to the east side of the channel is relatively undisturbed. Direct impacts may be runoff from streets into channel during rain events.

Specific Site Characteristics: Pond Restored (PR)

a) Location of site: The datalogger is located at the middle levee breach between Pond Eleven and Pond Ten, which is part of the South San Diego Bay Coastal Wetland Restoration and Enhancement Project; 32° 35' 45.9", 117° 7' 5.5" W.

b) Elevation of sonde's depth port:

February 25, 2015 - January 7, 2021: -0.310m NAVD88, approximately 0.5m above the channel bottom.

January 7, 2021 - November 30, 2021: -0.381m NAVD88, approximately 0.5m above the channel bottom.

November 30, 2021 – present: -0.397m NAVD88, approximately 0.5m above the channel bottom.

c) Water depth: 0.5m to 3.2m.

d) Channel width: approximately 40m.

e) Bottom type: silt and clay.

f) Tidal Exchange (extremes): approximately -2ft to +7ft MLLW.

g) Salinity: 4ppt (extreme rain event) to 39 ppt (average of 35ppt)

Specific Site Characteristics: South Bay (SB)

a) Location of site: The datalogger is located at the mouth of Otay River where it flows into San Diego Bay; 32° 36' 3.6" N, 117° 06' 57.0" W.

b) Elevation of sonde's depth port:

February 25, 2015 – January 7, 2021: -0.379m NAVD88, approximately 0.5m above the channel bottom.

January 7, 2021 – December 1, 2021: -0.471m NAVD88 (EXO2) or -0.350m NAVD88 (EXO3), approximately 0.5m above the channel bottom.

December 1, 2021 – present: -0.629m NAVD88, approximately 0.5m above the channel bottom.

c) Depth: approximately 0.7m to 3.2m

d) Channel width: approximately 25m

e) Bottom type: silt and clay.

f) Tidal Exchange (extremes): approximately -2ft to +7ft MLLW.

g) Salinity: 4ppt (extreme rain event) to 39 ppt (average of 35ppt)

SWMP Station Timeline

Station Code	SWMP Status	Station Name	Location	Active Dates	Reason Decommissioned	Notes
BR (tjrbrwq)	P	Boca Rio	32° 33' 33.70 N, 117° 7' 44.30 W	12/23/2004 15:30	NA	NA
OS (tjroswq)	P	Oneonta Slough	32° 34' 6.00 N, 117° 7' 52.60 W	01/01/1996 00:00 -	NA	NA
PR (tjrprwq)	P	Pond Eleven Restored	32° 35' 45.90 N, 117° 07' 5.59 W	02/16/2012 11:00	NA	NA
SB (tjrswq)	P	South Bay	32° 36' 3.60 N, 117° 6' 57.00 W	01/02/2008 00:00 -	NA	NA
MM (tjrmmwq)	P	Model Marsh	32° 32' 52.08 N, 117° 7' 22.80 W	10/01/2000 00:00 - 01/17/2008 00:00	Heavy sedimentation compromised the station	
PE (tjrpewq)	P	Pond Eleven	32° 36' 3.54 N, 117° 06' 58.46 W	07/25/2008 00:00 - 09/29/2010 00:00	Deployments at this site were interrupted due to an extensive Restoration project	Restoration project was concluded in October 2011. Datalogger was relocated and renamed – Pond Eleven Restored, and deployments resumed in January 2012.
RC (tjrrewq)	P	River Channel	32° 33' 28.08 N, 117° 6' 21.96 W	08/01/2002 00:00 - 11/11/2004 14:00	Heavy sedimentation compromised the station	Replaced by Boca Rio site

TL (tjrtlwq)	P	Tidal Linkage	32° 34' 27.84 N, 117° 7' 37.92 W	05/01/1997 00:00 - 10/08/2007 00:00	Heavy sedimentation compromised the station	
-----------------	---	---------------	---	--	--	--

6) Data collection period –

Boca Rio (since 2004)

Deploy Date	Deploy Time	Retrieve Date	Retrieve Time
12/6/2022	12:30	1/18/2023	15:00
1/18/2023	15:15	2/14/2023	12:15
2/14/2023	12:30	3/14/2023	12:15
3/14/2023	12:30	4/19/2023	13:45
4/19/2023	14:00	5/16/2023	12:30
5/16/2023	12:45	6/13/2023	11:00
6/13/2023	11:15	7/11/2023	12:00
7/11/2023	12:15	8/15/2023	14:00
8/15/2023	14:15	9/27/2023	13:45
9/27/2023	14:15	11/21/2023	12:15
11/21/2023	12:30	12/19/2023	10:15
12/19/2023	10:30	1/24/2024	13:30

Oneonta Slough (since 1996)

Deploy Date	Deploy Time	Retrieve Date	Retrieve Time
12/7/2022	15:45	1/18/2023	15:00
1/18/2023	15:15	2/14/2023	11:30
2/14/2023	11:45	3/14/2023	13:15
3/14/2023	13:30	4/19/2023	12:30
4/19/2023	12:45	5/17/2023	14:30
5/17/2023	14:45	6/14/2023	13:30
6/14/2023	13:45	7/11/2023	13:00
7/11/2023	13:15	8/15/2023	11:00
8/15/2023	11:15	9/12/2023	11:30
9/12/2023	11:45	10/10/2023	12:30
10/10/2023	12:45	11/21/2023	9:15
11/21/2023	9:30	12/19/2023	12:00
12/19/2023	12:15	1/23/2024	13:00

Pond Restored (since 2012)

Deploy Date	Deploy Time	Retrieve Date	Retrieve Time
12/6/2022	13:15	1/18/2023	13:30
1/18/2023	13:45	2/15/2023	14:45
2/15/2023	15:00	3/16/2023	10:30
3/16/2023	10:45	4/18/2023	12:00
4/18/2023	12:15	5/17/2023	12:30
5/17/2023	12:45	6/14/2023	11:15
6/14/2023	11:30	7/11/2023	9:15
7/11/2023	9:30	8/16/2023	12:45
8/16/2023	13:00	9/13/2023	12:15
9/13/2023	12:30	10/11/2023	12:30
10/11/2023	12:45	11/21/2023	10:45
11/21/2023	11:00	12/19/2023	10:45
12/19/2023	11:00	1/24/2024	13:30

South Bay (since 2008)

Deploy Date	Deploy Time	Retrieve Date	Retrieve Time
12/6/2022	12:30	1/18/2023	12:00
1/18/2023	12:15	2/15/2023	15:15
2/15/2023	15:30	3/16/2023	10:00
3/16/2023	10:15	4/18/2023	11:30
4/18/2023	11:45	5/17/2023	12:00
5/17/2023	12:15	6/14/2023	10:45
6/14/2023	11:00	7/11/2023	9:45
7/11/2023	10:00	8/16/2023	12:00
8/16/2023	12:15	9/13/2023	11:30
9/13/2023	11:45	10/11/2023	11:45
10/11/2023	12:00	11/21/2023	10:15
11/21/2023	10:30	12/19/2023	11:15
12/19/2023	11:30	1/24/2024	12:45

7) Distribution –

NOAA retains the right to analyze, synthesize and publish summaries of the NERRS System-wide Monitoring Program data. The NERRS retains the right to be fully credited for having collected and processed the data. Following academic courtesy standards, the NERR site where the data were collected should be contacted and fully acknowledged in any subsequent publications in which any part of the data are used. The data set enclosed within this package/transmission is only as good as the quality assurance and quality control procedures outlined by the enclosed metadata reporting statement. The user bears all responsibility for its subsequent use/misuse in any further analyses or comparisons. The Federal government does not assume liability to the Recipient or third persons, nor will the Federal government reimburse or indemnify the Recipient for its liability due to any losses resulting in any way from the use of this data.

Requested citation format:

NOAA National Estuarine Research Reserve System (NERRS). System-wide Monitoring Program. Data accessed from the NOAA NERRS Centralized Data Management Office website: <http://www.nerrdata.org/>; accessed 12 October 2023.

Also include the following excerpt in the metadata which will address how and where the data can be obtained.

NERR water quality data and metadata can be obtained from the Research Coordinator at the individual NERR site (please see Principal Investigators and Contact Persons), from the Data Manager at the Centralized Data Management Office (please see personnel directory under the general information link on the CDMO home page) and online at the CDMO home page www.nerrdata.org. Data are available in comma delimited format.

8) Associated researchers and projects –

The research program at the TRNERR focuses on adaptive approaches to wetlands management, which involves coupling scientific investigation with management action. One focal area of research continues to be adaptive restoration, and the TRNERR has a long history of science-based restoration efforts. These programs incorporate descriptive and experimental approaches to investigate biotic and abiotic responses to marsh restoration, including ways to better achieve desired ecosystem responses. Two SWMP sites, based in South San Diego Bay, are associated with planned restoration of salt ponds in that area. Another active area of research is invasive species ecology and management. Although estuaries are typically invaded by a broad suite of species from many habitat types, current research is focusing on terrestrial and riparian invaders able to cross ecotones and invade salt marsh habitats. Researchers at the TRNERR are investigating mechanisms of invasions, impacts of invaders, and ecosystem recovery after exotic species control.

NERR SWMP water quality and weather data are used in a variety of reserve-based and external research and education programs. Water quality data from the Tijuana River, which rarely experienced mouth closure prior to 2016, provided an interesting contrast to data from other regional systems, which experience frequent closure events. During the 2016 El Niño and until recently, the river mouth experienced few closures, and therefore the water quality and nutrient data were crucial to detect the imminent closures as well as to identify the effects on the system. Besides the importance of the SWMP data for Research and Stewardship purposes, SWMP water quality data are incorporated into a high school curriculum developed at the reserve, serving as a great tool for the Education and Outreach programs.

Tier 1 nutrient sampling is being conducted monthly at all water quality datalogger stations. NERR SWMP meteorological data is collected at 15-minute intervals at 1 station which is located near the former Tidal Linkage water quality station. These data are available at www.nerrdata.org.

II. Physical Structure Descriptors

9) Sensor specifications –

TJR NERR deployed YSI EXO2 and YSI EXO3 sondes with depth, temperature/conductivity, Optical DO, pH and turbidity probes. At one of the sites (Oneonta Slough) Chlorophyll sensors were deployed.

YSI EXO Sonde:

Parameter: Temperature

Units: Celsius (C)

Sensor Type: Wiped probe; Thermistor

Model#: 599827

Range: -5 to 50 C

Accuracy: ± 0.2 C
Resolution: 0.001 C

Parameter: Conductivity
Units: milli-Siemens per cm (mS/cm)
Sensor Type: Wiped probe; 4-electrode cell with autoranging
Model#: 599827
Range: 0 to 100 mS/cm
Accuracy: $\pm 1\%$ of the reading or 0.002 mS/cm, whichever is greater
Resolution: 0.0001 to 0.01 mS/cm (range dependent)

Parameter: Salinity
Units: practical salinity units (psu)/parts per thousand (ppt)
Model#: 599827
Sensor Type: Wiped probe; Calculated from conductivity and temperature
Range: 0 to 70 ppt
Accuracy: $\pm 2\%$ of the reading or 0.2 ppt, whichever is greater
Resolution: 0.01 psu

Parameter: Dissolved Oxygen % saturation
Sensor Type: Optical probe w/ mechanical cleaning
Model#: 599100-01
Range: 0 to 500% air saturation
Accuracy: 0-200% air saturation: $\pm 1\%$ of the reading or 1% air saturation, whichever is greater 200-500% air saturation: $\pm 5\%$ of reading
Resolution: 0.1% air saturation

Parameter: Dissolved Oxygen mg/L (Calculated from % air saturation, temperature, and salinity)
Units: milligrams/Liter (mg/L)
Sensor Type: Optical probe w/ mechanical cleaning
Model#: 599100-01
Range: 0 to 50 mg/L
Accuracy: 0-20 mg/L: ± 0.1 mg/L or 1% of the reading, whichever is greater
20 to 50 mg/L: $\pm 5\%$ of the reading
Resolution: 0.01 mg/L

Parameter: Non-vented Level - Shallow (Depth)
Units: feet or meters (ft or m)
Sensor Type: Stainless steel strain gauge
Range: 0 to 33 ft (10 m)
Accuracy: ± 0.013 ft (0.004 m)
Resolution: 0.001 ft (0.001 m)

Parameter: pH
Units: pH units
Sensor Type: Glass combination electrode
Model#: 599701(guarded) or 599702(wiped)
Range: 0 to 14 units
Accuracy: ± 0.1 units within $\pm 10^\circ$ of calibration temperature, ± 0.2 units for entire temperature range
Resolution: 0.01 units

Parameter: Turbidity
Units: formazin nephelometric units (FNU)
Sensor Type: Optical, 90 degree scatter

Model#: 599101-01

Range: 0 to 4000 FNU

Accuracy: 0 to 999 FNU: 0.3 FNU or +/-2% of reading (whichever is greater); 1000 to 4000 FNU +/-5% of reading

Resolution: 0 to 999 FNU: 0.01 FNU, 1000 to 4000 FNU: 0.1 FNU

Parameter: Chlorophyll

Units: micrograms/Liter

Sensor Type: Optical probe

Model#: 599102-01

Range: 0 to 400 ug/Liter

Accuracy: Dependent on methodology

Resolution: 0.1 ug/L chl a, 0.1% FS

Depth Qualifier:

The NERR System-Wide Monitoring Program utilizes YSI data sondes that can be equipped with either vented or non-vented depth/level sensors. Readings for both vented and non-vented sensors are automatically compensated for water density change due to variations in temperature and salinity; but for all non-vented depth measurements, changes in atmospheric pressure between calibrations appear as changes in water depth. The error is equal to approximately 1.02 cm for every 1 millibar change in atmospheric pressure, and is eliminated for vented sensors because they are vented to the atmosphere throughout the deployment time interval.

Beginning in 2006, NERR SWMP standard calibration protocol calls for all non-vented depth sensors to read 0 meters at a (local) barometric pressure of 1013.25 mb (760 mm/hg). To achieve this, each site calibrates their depth sensor with a depth offset number, which is calculated using the actual atmospheric pressure at the time of calibration and the equation provided in the SWMP calibration sheet or digital calibration log. This offset procedure standardizes each depth calibration for the entire NERR System. If accurate atmospheric pressure data are available, non-vented sensor depth measurements at any NERR can be corrected.

In 2010, the CDMO began automatically correcting Depth/Level data for changes in barometric pressure as measured by the reserve's associated meteorological station during data ingestion. These corrected Depth/Level data are reported as cDepth and cLevel, and are assigned QAQC flags and codes based on QAQC protocols. Please see sections 11 and 12 for QAQC flag and code definitions.

NOTE: older Depth data cannot be corrected without verifying that the depth offset was in place and whether a vented or non-vented depth sensor was in use. No SWMP data prior to 2006 can be corrected using this method. The following equation is used for corrected Depth/Level data provided by the CDMO beginning in 2010:
$$((1013-BP)*0.0102)+\text{Depth/Level} = \text{cDepth/cLevel}.$$

Salinity Units Qualifier:

In 2013, EXO sondes were approved for SWMP use and began to be utilized by reserves. While the 6600 series sondes report salinity in parts per thousand (ppt) units, the EXO sondes report practical salinity units (psu). These units are essentially the same and for SWMP purposes are understood to be equivalent, however psu is considered the more appropriate designation. Moving forward the NERR System will assign psu salinity units for all data regardless of sonde type.

Turbidity Qualifier:

In 2013, EXO sondes were approved for SWMP use and began to be utilized by reserves. While the 6600 series sondes report turbidity in nephelometric turbidity units (NTU), the EXO sondes use formazin nephelometric units (FNU). These units are essentially the same but indicate a difference in sensor methodology, for SWMP purposes they will be considered equivalent. Moving forward, the

NERR System will use FNU/NTU as the designated units for all turbidity data regardless of sonde type. If turbidity units and sensor methodology are of concern, please see the Sensor Specifications portion of the metadata.

Chlorophyll Fluorescence Disclaimer:

YSI chlorophyll sensors (6025 or 599102-01) are designed to serve as a proxy for chlorophyll concentrations in the field for monitoring applications and complement traditional lab extraction methods; therefore, there are accuracy limitations associated with the data that are detailed in the YSI manual including interference from other fluorescent species, differences in calibration method, and effects of cell structure, particle size, organism type, temperature, and light on sensor measurements.

10) Coded variable definitions –

Sampling station:	Sampling site code:	Station code:
Boca Rio	BR	tjrbrwq
Oneonta Slough	OS	tjroswq
Pond Restored	PR	tjrprwq
South Bay	SB	tjrswbq

11) QAQC flag definitions –

QAQC flags provide documentation of the data and are applied to individual data points by insertion into the parameter's associated flag column (header preceded by an F_). During primary automated QAQC (performed by the CDMO), -5, -4, and -2 flags are applied automatically to indicate data that is missing and above or below sensor range. All remaining data are then flagged 0, passing initial QAQC checks. During secondary and tertiary QAQC 1, -3, and 5 flags may be used to note data as suspect, rejected due to QAQC, or corrected.

- 5 Outside High Sensor Range
- 4 Outside Low Sensor Range
- 3 Data Rejected due to QAQC
- 2 Missing Data
- 1 Optional SWMP Supported Parameter
- 0 Data Passed Initial QAQC Checks
- 1 Suspect Data
- 2 *Open - reserved for later flag*
- 3 Calculated data: non-vented depth/level sensor correction for changes in barometric pressure
- 4 Historical Data: Pre-Auto QAQC
- 5 Corrected Data

12) QAQC code definitions –

QAQC codes are used in conjunction with QAQC flags to provide further documentation of the data and are also applied by insertion into the associated flag column. There are three (3) different code categories, general, sensor, and comment. General errors document general problems with the deployment or YSI datasonde, sensor errors are sensor specific, and comment codes are used to further document conditions or a problem with the data. Only one general or sensor error and one comment code can be applied to a particular data point, but some comment codes (marked with an * below) can be applied to the entire record in the F_Record column.

General Errors

- GIC No instrument deployed due to ice
- GIM Instrument malfunction

GIT	Instrument recording error; recovered telemetry data
GMC	No instrument deployed due to maintenance/calibration
GNF	Deployment tube clogged / no flow
GOW	Out of water event
GPF	Power failure / low battery
GQR	Data rejected due to QA/QC checks
GSM	See metadata

Corrected Depth/Level Data Codes

GCC	Calculated with data that were corrected during QA/QC
GCM	Calculated value could not be determined due to missing data
GCR	Calculated value could not be determined due to rejected data
GCS	Calculated value suspect due to questionable data
GCU	Calculated value could not be determined due to unavailable data

Sensor Errors

SBO	Blocked optic
SCF	Conductivity sensor failure
SCS	Chlorophyll spike
SDF	Depth port frozen
SDG	Suspect due to sensor diagnostics
SDO	DO suspect
SDP	DO membrane puncture
SIC	Incorrect calibration / contaminated standard
SNV	Negative value
SOW	Sensor out of water
SPC	Post calibration out of range
SQR	Data rejected due to QAQC checks
SSD	Sensor drift
SSM	Sensor malfunction
SSR	Sensor removed / not deployed
STF	Catastrophic temperature sensor failure
STS	Turbidity spike
SWM	Wiper malfunction / loss

Comments

CAB*	Algal bloom
CAF	Acceptable calibration/accuracy error of sensor
CAP	Depth sensor in water, affected by atmospheric pressure
CBF	Biofouling
CCU	Cause unknown
CDA*	DO hypoxia (<3 mg/L)
CDB*	Disturbed bottom
CDF	Data appear to fit conditions
CFK*	Fish kill
CIP*	Surface ice present at sample station
CLT*	Low tide
CMC*	In field maintenance/cleaning
CMD*	Mud in probe guard
CND	New deployment begins
CRE*	Significant rain event
CSM*	See metadata
CTS	Turbidity spike
CVT*	Possible vandalism/tampering
CWD*	Data collected at wrong depth

CWE* Significant weather event

13) Post deployment information –

Boca Rio

Deploy Date	Deploy Time	Retrieve Date	Retrieve Time	Sonde Model (Nickname)	pH Model	ODO Model	Turb Model	Cond Model	EXO Model Number
1/18/2023	15:15	2/14/2023	12:15	EXO2 (BR)	599702	599100-1	599101-01	599827	599090-01
2/14/2023	12:30	3/14/2023	12:15	EXO2 (BR2 exo)	599702	599100-01	599101-01	599827	599090-01
3/14/2023	12:30	4/19/2023	13:45	EXO2 (BR exo)	599702	599100-01	599101-01	599827	599090-01
4/19/2023	14:00	5/16/2023	12:30	EXO2 (BR2 exo)	599702	599100-01	599101-01	599827	599090-01
5/16/2023	12:30	6/13/2023	11:00	EXO2 (BR exo)	599702	599100-01	599101-01	599827	599090-01
6/13/2023	11:15	7/11/2023	12:00	EXO2 (BR2 exo)	599702	599100-01	599101-01	599827	599090-01
7/11/2023	12:15	8/15/2023	14:00	EXO2 (BR exo)	599702	599100-01	599101-01	599827	599090-01
8/15/2023	14:15	9/27/2023	13:45	EXO2 (BR2 exo)	599702	599100-01	599101-01	599827	599090-01
9/27/2023	14:15	10/9/2023	12:30	EXO2 (BR exo)	599702	599100-01	599101-01	599827	599090-01
10/9/2023	12:45	11/21/2023	12:15	EXO2 (BR2 exo)	599702	599100-01	599101-01	599827	599090-01
11/21/2023	12:30	12/19/2023	10:15	EXO2 (BR exo)	599702	599100-01	599101-01	599827	599090-01
12/19/2023	10:30	1/24/2024	13:30	EXO2 (BR2 exo)	599702	599100-01	599101-01	599827	599090-01
Deploy Date	Sonde Nickn	SpCond	ODO1	ODO2	pH7	pH10	Turb	Turb	Level
1/18/2023	BR	53.18(50.0)	95.8	94.9	7.06	10.05	0.38(0.0)	128.68(124.0)	-0.051(-0.005)
2/14/2023	BR2 exo	48.26(50.0)	98.1	96.7	7	10.05	0.71(0.0)	121.54(124.0)	-0.037(-0.032)
3/14/2023	BR exo	51.33(50.0)	98.6	98.6	7.03	10	0.31(0.0)	130.22(124.0)	0.004(-0.012)
4/19/2023	BR2 exo	47.69(50.0)	98.1	97.8	6.97	9.95	0.05(0.0)	125.25(124.0)	-0.069(-0.07)
5/16/2023	BR exo	50.99(50.0)	97.1	97.3	6.97	10.03	0.35(0.0)	132.02(124.0)	-0.071(-0.073)
6/13/2023	BR2 exo	50.73(50.0)	101.1	101.1	7.12	10.13	0.03(0.0)	124.51(124.0)	-0.052(-0.053)
7/11/2023	BR exo	50.16(50.0)	98	97.9	7.1	10.1	0.0(0.0)	127.84(124.0)	-0.086(-0.079)
8/15/2023	BR2 exo	50.52(50.0)	99.8	100	7.14	10.09	0.0(0.0)	122.93(124.0)	-0.12(-0.099)
9/27/2023	BR exo	50.55(50.0)	97.3	97.3	7.04	10.04	0.1(0.0)	124.65(124.0)	-0.143(-0.145)
10/9/2023	BR2 exo	54.43(50.0)	101.5	101.4	7.1	10.05	0.06(0.0)	133.64(124.0)	(-0.03)
11/21/2023	BR exo	50.41(50.0)	94	94	7.02	10.01	-0.04(0.0)	122.86(124.0)	-0.086(-0.068)
12/19/2023	BR2 exo	51.54(50.0)	87.2	87.1	6.92	9.85	0.19(0.0)	128.64(124.0)	0.009(0.025)

Oneonta Slough

Deploy Date	Deploy Time	Retrieve Date	Retrieve Time	Sonde Model (Nickname)	pH Model	ODO Model	Turb Model	Cond Model	EXO Model Number	Chloro Model	
1/18/2023	15:15	2/14/2023	11:30	EXO2 (OS_EXO)	599702	599100-01	599101-01	599827	599090-01	599103-01	
2/14/2023	11:45	3/14/2023	13:15	EXO2 (OS2 exo)	599702	599100-1	599101-1	599827	599090-01	599103-01	
3/14/2023	13:30	4/19/2023	12:30	EXO2 (OS exo)	599702	599100-1	599101-01	599827	599090-01	599103-01	
4/19/2023	12:45	5/17/2023	14:30	EXO2 (OS2 exo)	599702	599100-1	599101-1	599827	599090-01	599103-01	
5/17/2023	14:45	6/14/2023	13:30	EXO2 (OS exo)	599702	599100-1	599101-01	599827	599090-01	599103-01	
6/14/2023	13:45	7/11/2023	13:00	EXO2 (OS2 exo)	599702	599100-1	599101-1	599827	599090-01	599103-01	
7/11/2023	13:15	8/15/2023	11:00	EXO2 (OS exo)	599702	599100-1	599101-01	599827	599090-01	599103-01	
8/15/2023	11:15	9/12/2023	11:30	EXO2 (OS2 exo)	599702	599100-1	599101-1	599827	599090-01	599103-01	
9/12/2023	11:45	10/10/2023	12:30	EXO2 (OS exo)	599702	599100-1	599101-01	599827	599090-01	599103-01	
10/10/2023	12:45	11/21/2023	9:15	EXO2 (OS2 exo)	599702	599100-1	599101-1	599827	599090-01	599103-01	
11/21/2023	9:30	12/19/2023	12:00	EXO2 (OS exo)	599702	599100-1	599101-01	599827	599090-01	599103-01	
12/19/2023	12:15	1/23/2024	13:00	EXO2 (OS2 exo)	599702	599100-1	599101-1	599827	599090-01	599103-01	
Deploy Date	Sonde Nickname	SpCond	ODO1	ODO2	pH7	pH10	Turb	Turb	Level	CHL(0)	CHL(118)
1/18/2023	OS_EXO	44.46(50.0)	102.1	100.7	7.14	10.1	1.71(0.0)	128.39(124.0)	0.241(0.0)	0.14	79.89
2/14/2023	OS2 exo	49.05(50.0)	98.1	96.7	6.94	10.07	0.4(0.0)	123.2(124.0)	0.284(0.286)	1.84	89.4
3/14/2023	OS exo	50.75(50.0)	94.6	93.9	7.18	10.1	0.35(0.0)	125.89(124.0)	0.287(0.298)	-0.02	68.05
4/19/2023	OS2 exo	47.14(50.0)	97.6	97.3	7.07	10.1	-0.04(0.0)	128.27(124.0)	0.257(0.266)	0.05	70.32
5/17/2023	OS exo	50.39(50.0)	92.3	92.8	6.96	9.96	-0.04(0.0)	123.17(124.0)	0.259(0.262)	-0.09	68.94
6/14/2023	OS2 exo	50.28(50.0)	75.1	75.1	7.1	10.08	0.07(0.0)	124.98(124.0)	0.275(0.275)	-0.15	70.9
7/11/2023	OS exo	50.21(50.0)	99.5	99.6	7.04	10.1	0.08(0.0)	118.7(124.0)	0.257(0.259)	-0.09	58.9
8/15/2023	OS2 exo	50.24(50.0)	100.4	100.5	7.13	10.08	0.0(0.0)	124.42(124.0)	0.274(0.274)	0.21	73.54
9/12/2023	OS exo	50.84(50.0)	101.3	101.2	7.07	10.05	0.03(0.0)	132.35(124.0)	0.198(0.198)	0.34	76.3
10/10/2023	OS2 exo	51.28(50.0)	95	95.2	7.15	10.16	0.1(0.0)	126.52(124.0)	0.312(0.311)	-0.28	63.6
11/21/2023	OS exo	50.76(50.0)	105.6	105.1	7.18	10.12	0.03(0.0)	127.61(124.0)	0.25(0.271)	-0.12	70.56
12/19/2023	OS2 exo	51.08(50.0)	95.2	95.3	7.15	10.14	0.19(0.0)	129.51(124.0)	0.252(0.279)	0.08	68.99

Pond Restored

Deploy Date	Deploy Time	Retrieve Date	Retrieve Time	Sonde Model (Nickname)	pH Model	ODO Model	Turb Model	Cond Model	EXO Model Number
1/18/2023	13:30	2/15/2023	14:45	EXO3 ()	599702	599100-1	599101-1	599827	599090-01
2/15/2023	14:45	3/16/2023	10:30	EXO3 (PR exo)	599702	599100-1	599101-1	599827	599090-1
3/16/2023	10:30	4/18/2023	13:00	EXO3 (PR2 exo)	599702	599100-1	599101-1	599827	599090-1
4/18/2023	13:00	5/17/2023	12:30	EXO3 (PR exo)	599702	599100-1	599101-1	599827	599090-1
5/17/2023	12:30	6/14/2023	11:15	EXO3 (PR2 exo)	599702	599100-1	599101-1	599827	599090-1
6/14/2023	11:15	7/11/2023	9:15	EXO3 (PR exo)	599702	599100-1	599101-1	599827	599090-1
7/11/2023	9:15	8/16/2023	12:45	EXO3 (PR2 exo)	599702	599100-1	599101-1	599827	599090-1
8/16/2023	12:45	9/13/2023	12:15	EXO3 (PR exo)	599702	599100-1	599101-1	599827	599090-1
9/13/2023	12:15	10/11/2023	12:30	EXO3 (PR2 exo)	599702	599100-1	599101-01	599827	599090-1
10/11/2023	12:30	11/21/2023	10:45	EXO2 (Uca)	599702	599100-1	599101-1	599827	599090-1
11/21/2023	10:45	12/19/2023	13:45	EXO3 (PR2 exo)	599702	599100-1	599101-01	599827	599090-1
12/19/2023	11:00	1/24/2024	13:30	EXO3 (PR exo)	599702	599100-1	599101-1	599827	599090-1
Deploy Date	Sonde Nickname	SpCond	ODO1	ODO2	pH7	pH10	Turb	Turb	Level
1/18/2023		43.42(50.0)	99.9	100	7.01	10	0.54(0.0)	131.97(124.0)	-0.328(-0.321)
2/15/2023	PR exo	48.23(50.0)	100.8	99.8	7.02	10.07	0.05(0.0)	118.23(124.0)	-0.369(-0.367)
3/16/2023	PR2 exo	50.76(50.0)	100.4	100.3	7.03	10.01	0.17(0.0)	127.61(124.0)	-0.365(-0.358)
4/18/2023	PR exo	48.37(50.0)	100.9	100.9	6.97	10	0.1(0.0)	124.66(124.0)	-0.457(-0.362)
5/17/2023	PR2 exo	50.13(50.0)	96.8	96.6	6.94	10.02	0.67(0.0)	122.72(124.0)	-0.396(-0.398)
6/14/2023	PR exo	50.33(50.0)	84.2	84.5	6.95	9.95	4.31(0.0)	110.91(124.0)	-0.386(-0.385)
7/11/2023	PR2 exo	50.47(50.0)	101.2	101.3	6.98	10.04	0.06(0.0)	125.34(124.0)	-0.431(-0.431)
8/16/2023	PR exo	45.79(50.0)	101	109.9	7.16	9.96	-0.01(0.0)	123.97(124.0)	-0.398(-0.401)
9/13/2023	PR2 exo	51.07(50.0)	101.2	101.2	7.06	10.02	0.13(0.0)	126.52(124.0)	-0.418(-0.419)
10/11/2023	Uca	47.32(50.0)	104.1	104.1	7.12	10.07	0.01(0.0)	136.64(124.0)	-0.333(-0.329)
11/21/2023	PR2 exo	50.74(50.0)	100.4	100.3	7.01	10	0.08(0.0)	125.09(124.0)	-0.417(-0.398)
12/19/2023	PR exo	50.06(50.0)	99.5	99.5	7.08	10.04	0.11(0.0)	126.92(124.0)	-0.331(-0.191)

South Bay

Deploy Date	Deploy Time	Retrieve Date	Retrieve Time	Sonde Model (Nickname)	pH Model	ODO Model	Turb Model	Cond Model	EXO Model Number
1/18/2023	12:15	2/15/2023	15:15	EXO3 (SB-2 exo)	599702	599100-1	599101-1	599827	599090-01
2/15/2023	15:15	3/16/2023	10:00	EXO3 (SB exo)	599702	599100-1	599101-1	599827	599090-1
3/16/2023	10:00	4/18/2023	12:30	EXO3 (SB2 exo)	599702	599100-1	599101-1	599827	599090-1
4/18/2023	12:30	5/17/2023	12:00	EXO3 (SB exo)	599702	599100-1	599101-1	599827	599090-1
5/17/2023	12:00	6/14/2023	10:45	EXO3 (SB2 exo)	599702	599100-1	599101-1	599827	599090-1
6/14/2023	10:45	7/11/2023	9:45	EXO2 (Uca)	599702	599100-1	599101-1	599827	599090-1
7/11/2023	9:45	8/16/2023	12:00	EXO3 (SB2 exo)	599702	599100-1	599101-1	599827	599090-1
8/16/2023	12:00	9/13/2023	11:30	EXO3 (SB exo)	599702	599100-1	599101-1	599827	599090-1
9/13/2023	11:30	10/11/2023	11:45	EXO3 (SB2 exo)	599702	599100-1	599101-01	599827	599090-1
10/11/2023	11:45	11/21/2023	10:15	EXO3 (SB exo)	599702	599100-1	599101-1	599827	599090-1
11/21/2023	10:30	12/19/2023	14:15	EXO3 (SB2 exo)	599702	599100-1	599101-01	599827	599090-1
12/19/2023	14:15	1/24/2024	12:45	EXO3 (SB exo)	599702	599100-1	599101-1	599827	599090-1
Deploy Date	Sonde Nickname	SpCond	ODO1	ODO2	pH7	pH10	Turb	Turb	Level
1/18/2023	SB-2 exo	52.43(50.0)	99.6	99.7	7.04	10.01	0.9(0.0)	134.4(124.0)	-0.552(-0.545)
2/15/2023	SB exo	48.19(50.0)	100.3	100.3	7.04	10.03	0.05(0.0)	122.28(124.0)	0.109(-0.603)
3/16/2023	SB2 exo	50.27(50.0)	100	100	7.06	10.08	0.03(0.0)	126.52(124.0)	-0.614(-0.618)
4/18/2023	SB exo	51.34(50.0)	99.9	99.9	7.01	10.01	0.08(0.0)	133.25(124.0)	-0.315(-0.626)
5/17/2023	SB2 exo	50.2(50.0)	98	98.2	6.94	10	0.29(0.0)	124.82(124.0)	-0.636(-0.633)
6/14/2023	Uca	50.48(50.0)	98.1	98.4	7.02	10.05	-0.23(0.0)	121.06(124.0)	-0.62(-0.621)
7/11/2023	SB2 exo	missing	101.4	101.4	7.04	10.07	0.11(0.0)	125.32(124.0)	-0.633(-0.662)
8/16/2023	SB exo	45.74(50.0)	101.4	101.4	7.08	10.04	-0.34(0.0)	125.71(124.0)	-0.629(-0.63)
9/13/2023	SB2 exo	51.07(50.0)	102.2	102.2	7.09	10.05	0.06(0.0)	126.62(124.0)	-0.649(-0.648)
10/11/2023	SB exo	46.39(50.0)	102.8	102.8	7.08	10.04	0.22(0.0)	130.88(124.0)	-0.522(-0.552)
11/21/2023	SB2 exo	50.67(50.0)	95.7	95.8	7.02	10	0.0(0.0)	128.52(124.0)	(-0.657)
12/19/2023	SB exo	-50	78.3	78.3	7.08	10.08	0.31(0.0)	125.0(124.0)	-0.56(-0.439)

14) Other remarks/notes –

Data are missing due to equipment or associated specific probes not being deployed, equipment failure, time of maintenance or calibration of equipment, or repair/replacement of a sampling station platform. Any NaNs in the dataset stand for “not a number” and are the result of low power, disconnected wires, or out of range readings. If additional information on missing data is needed, contact the Research Coordinator at the reserve submitting the data.

Precipitation

The rain data below were recorded by a tipping bucket in the TJR NERR Tidal Link meteorological station except where otherwise noted.

Date	.in
1/3	1.01 (manual gauge)
1/14 - 1/17	2.47
2/14	0.48 (manual gauge)
2/20	0.01 (manual gauge)
2/21	0.06
2/22	0.22
2/23 - 2/27	1.25
2/27 - 2/28	0.03
2/28 - 3/1	0.25
3/10 - 3/13	0.54
3/14 - 3/15	0.72
3/19 - 3/20	0.03
3/21 - 3/23	1.19
3/29 – 3/30	0.41
4/11 - 4/13	0.04
5/2 – 5/5	0.13 (manual gauge)
6/11 – 6/13	0.04
8/19 - 8/21	1.81
9/2	0.03
10/12	0.03
10/16	0.02 (manual gauge)
11/27	0.49
12/14	0.05 (manual gauge)
12/21	0.02 (manual gauge)
12/22	0.30 (manual gauge)
12/31 – (1/3/24)	0.54

Site specific issues

Two of our sites are considerably affected by transboundary flows of the Tijuana River that consist of partially treated and untreated wastewater. Data showing both recorded and nearly real-time daily discharge of the Tijuana River, measured by the U.S. International Boundary & Water Commission

(IBWC) stream gauge, can be obtained at the [IBWC water data dashboard](#). During such flows, staff is oriented to avoid contact with the water.

Boca Rio

01/01 – 01/18 The depth sensor did not record data due to DO sensor resetting and not allowing the sonde to sleep properly between logging measurements. It was verified with YSI and determined that this issue did not affect the data.

10/12 – 10/16 and 10/26 – 11/16 Level data is corrected using the barometric pressure (BP) data from our meteorological station. During this period the BP sensor presented some issues.

Both Boca Rio and Oneonta Slough stations recorded negative DO values in some of the deployments. Per protocol, these negative values exceed the sensor limits and therefore, must be rejected. However, some negative data seem reflective of the anoxic conditions that are frequently observed at these sites. Such conditions were more frequent during this first quarter due to a great amount of rain. The DO mg/l values were marked suspect when it was within the sensor range but the percentage was below the sensor range. Rejected data can be requested from the CDMO data download page by selecting the “All data” option.

Oneonta Slough

01/01 – 01/18 Chlorophyll started to be reported only on January 18, data is missing from the first of the month until then.

10/12 – 10/16 and 10/26 – 11/16 Level data is corrected using the barometric pressure (BP) data from our meteorological station. During this period the BP sensor presented some issues.

Both Boca Rio and Oneonta Slough stations recorded negative DO values in some of the deployments. Per protocol, these negative values exceed the sensor limits and therefore, must be rejected. However, some negative data seem reflective of the anoxic conditions that are frequently observed at these sites. Such conditions were more frequent during this first quarter due to a great amount of rain. The DO mg/l values were marked suspect when it was within the sensor range but the percentage was below the sensor range. Rejected data can be requested from the CDMO data download page by selecting the “All data” option.

Pond Restored

10/12 – 10/16 and 10/26 – 11/16 Level data is corrected using the barometric pressure (BP) data from our meteorological station. During this period the BP sensor presented some issues.

South Bay

10/12 – 10/16 and 10/26 – 11/16 Level data is corrected using the barometric pressure (BP) data from our meteorological station. During this period the BP sensor presented some issues.